

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082319\
 Data File : VW012151.D
 Acq On : 23 Aug 2019 23:16
 Operator : SY/VA
 Sample : K4474-14
 Misc : 6.50G/5ML/MSVOA W/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T1-2-K3-(0)

Quant Time: Aug 26 08:23:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	218432	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	367217	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	303631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	112601	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	143884	55.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.82%	
35) Dibromofluoromethane	7.88	113	111909	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
50) Toluene-d8	10.32	98	454024	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
62) 4-Bromofluorobenzene	12.62	95	136503	43.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.14%	
Target Compounds						
16) Acetone	4.14	43	21761	32.785	ug/l	97
20) Methylene Chloride	4.90	84	20346	8.108	ug/l	96
29) Tetrahydrofuran	7.54	42	4562	8.430	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082319\
Data File : VW012151.D
Acq On : 23 Aug 2019 23:16
Operator : SY/VA
Sample : K4474-14
Misc : 6.50G/5ML/MSVOA W/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T1-2-K3-(0)

Quant Time: Aug 26 08:23:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

