

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082319\
 Data File : VW012130.D
 Acq On : 23 Aug 2019 13:41
 Operator : SY/VA
 Sample : VSTDICV050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ICVVW082319

Quant Time: Aug 26 07:25:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	0.308	0.355	-15.3	111	0.00
3 P	Chloromethane	0.477	0.490	-2.7	105	0.00
4 C	Vinyl Chloride	0.584	0.615	-5.3#	104	0.00
5 T	Bromomethane	0.296	0.313	-5.7	102	0.00
6 T	Chloroethane	0.329	0.345	-4.9	102	0.00
7 T	Trichlorofluoromethane	0.293	0.324	-10.6	108	0.00
8 T	Diethyl Ether	0.278	0.280	-0.7	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.498	0.521	-4.6	103	0.00
10 T	Methyl Iodide	0.674	0.702	-4.2	101	0.00
11 T	Tert butyl alcohol	0.036	0.036	0.0	100	0.00
12 CM	1,1-Dichloroethene	0.502	0.528	-5.2#	104	0.00
13 T	Acrolein	0.034	0.031	8.8	101	0.00
14 T	Allyl chloride	1.072	1.138	-6.2	104	0.00
15 T	Acrylonitrile	0.138	0.138	0.0	101	0.00
16 T	Acetone	0.152	0.170	-11.8	101	0.00
17 T	Carbon Disulfide	1.516	1.670	-10.2	104	0.00
18 T	Methyl Acetate	0.376	0.360	4.3	98	0.00
19 T	Methyl tert-butyl Ether	0.811	0.835	-3.0	102	0.00
20 T	Methylene Chloride	0.574	0.558	2.8	102	0.00
21 T	trans-1,2-Dichloroethene	0.538	0.558	-3.7	102	0.00
22 T	Diisopropyl ether	1.987	2.016	-1.5	100	0.00
23 T	Vinyl Acetate	1.223	1.281	-4.7	100	0.00
24 P	1,1-Dichloroethane	1.082	1.088	-0.6	100	0.00
25 T	2-Butanone	0.206	0.214	-3.9	98	0.00
26 T	2,2-Dichloropropane	0.612	0.647	-5.7	107	0.00
27 T	cis-1,2-Dichloroethene	0.575	0.587	-2.1	100	0.00
28 T	Bromochloromethane	0.472	0.472	0.0	111	0.00
29 T	Tetrahydrofuran	0.124	0.128	-3.2	100	0.00
30 C	Chloroform	0.990	0.982	0.8#	100	0.00
31 T	Cyclohexane	1.071	1.067	0.4	102	0.00
32 T	1,1,1-Trichloroethane	0.783	0.812	-3.7	104	0.00
33 S	1,2-Dichloroethane-d4	0.594	0.552	7.1	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.298	0.282	5.4	98	0.00
36 T	1,1-Dichloropropene	0.517	0.547	-5.8	104	0.00
37 T	Ethyl Acetate	0.283	0.286	-1.1	100	0.00
38 T	Carbon Tetrachloride	0.447	0.470	-5.1	103	0.00
39 T	Methylcyclohexane	0.580	0.628	-8.3	104	0.00
40 TM	Benzene	1.411	1.438	-1.9	101	0.00
41 T	Methacrylonitrile	0.166	0.187	-12.7	111	0.00
42 TM	1,2-Dichloroethane	0.442	0.441	0.2	99	0.00
43 T	Isopropyl Acetate	0.524	0.537	-2.5	101	0.00
44 TM	Trichloroethene	0.343	0.352	-2.6	102	0.00
45 C	1,2-Dichloropropane	0.380	0.381	-0.3#	100	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082319\
 Data File : VW012130.D
 Acq On : 23 Aug 2019 13:41
 Operator : SY/VA
 Sample : VSTDICV050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ICVVW082319

Quant Time: Aug 26 07:25:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.169	0.170	-0.6	100	0.00
47 T	Bromodichloromethane	0.450	0.456	-1.3	99	0.00
48 T	Methyl methacrylate	0.250	0.274	-9.6	105	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	101	0.00
50 S	Toluene-d8	1.210	1.184	2.1	100	0.00
51 T	4-Methyl-2-Pentanone	0.267	0.275	-3.0	100	0.00
52 CM	Toluene	0.848	0.873	-2.9#	101	0.00
53 T	t-1,3-Dichloropropene	0.458	0.483	-5.5	101	0.00
54 T	cis-1,3-Dichloropropene	0.553	0.578	-4.5	101	0.00
55 T	1,1,2-Trichloroethane	0.239	0.240	-0.4	101	0.00
56 T	Ethyl methacrylate	0.348	0.368	-5.7	101	0.00
57 T	1,3-Dichloropropane	0.452	0.460	-1.8	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.179	0.178	0.6	105	0.00
59 T	2-Hexanone	0.189	0.203	-7.4	100	0.00
60 T	Dibromochloromethane	0.266	0.276	-3.8	100	0.00
61 T	1,2-Dibromoethane	0.223	0.228	-2.2	100	0.00
62 S	4-Bromofluorobenzene	0.427	0.409	4.2	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.324	0.334	-3.1	100	0.00
65 PM	Chlorobenzene	0.985	1.007	-2.2	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.343	0.355	-3.5	100	0.00
67 C	Ethyl Benzene	1.897	1.983	-4.5#	103	0.00
68 T	m/p-Xylenes	0.683	0.719	-5.3	103	0.00
69 T	o-Xylene	0.632	0.670	-6.0	102	0.00
70 T	Styrene	1.094	1.146	-4.8	100	0.00
71 P	Bromoform	0.171	0.176	-2.9	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.747	3.954	-5.5	102	0.00
74 T	N-amyl acetate	1.132	1.213	-7.2	101	0.00
75 P	1,1,2,2-Tetrachloroethane	0.687	0.689	-0.3	99	0.00
76 T	1,2,3-Trichloropropane	0.487	0.496	-1.8	101	0.00
77 T	Bromobenzene	0.811	0.847	-4.4	103	0.00
78 T	n-propylbenzene	4.574	4.844	-5.9	102	0.00
79 T	2-Chlorotoluene	2.597	2.730	-5.1	103	0.00
80 T	1,3,5-Trimethylbenzene	3.131	3.319	-6.0	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.222	0.243	-9.5	103	0.00
82 T	4-Chlorotoluene	2.719	2.818	-3.6	101	0.00
83 T	tert-Butylbenzene	2.652	2.817	-6.2	103	0.00
84 T	1,2,4-Trimethylbenzene	3.145	3.305	-5.1	102	0.00
85 T	sec-Butylbenzene	3.799	4.048	-6.6	104	0.00
86 T	p-Isopropyltoluene	3.393	3.662	-7.9	104	0.00
87 T	1,3-Dichlorobenzene	1.585	1.629	-2.8	103	0.00
88 T	1,4-Dichlorobenzene	1.574	1.624	-3.2	102	0.00
89 T	n-Butylbenzene	3.374	3.693	-9.5	104	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082319\
Data File : VW012130.D
Acq On : 23 Aug 2019 13:41
Operator : SY/VA
Sample : VSTDICV050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ICVW082319

Quant Time: Aug 26 07:25:15 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.595	0.635	-6.7	102	0.00
91 T	1,2-Dichlorobenzene	1.391	1.442	-3.7	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.116	0.119	-2.6	103	0.00
93 T	1,2,4-Trichlorobenzene	0.961	1.061	-10.4	105	0.00
94 T	Hexachlorobutadiene	0.635	0.684	-7.7	107	0.00
95 T	Naphthalene	1.683	1.913	-13.7	104	0.00
96 T	1,2,3-Trichlorobenzene	0.839	0.928	-10.6	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6