

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082323\
 Data File : VW026841.D
 Acq On : 24 Aug 2023 01:45
 Operator : SY/MD
 Sample : 04113-03
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EZYH6

Quant Time: Aug 24 04:50:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Aug 23 01:29:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	444154	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	315764	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	88214	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	79967	10.952	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	43.800%		
7) Chloroethane-d5	2.893	69	79669	14.674	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	58.680%		
11) 1,1-Dichloroethene-d2	4.027	65	49339	13.739	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	54.960%		
21) 2-Butanone-d5	7.081	46	103695	49.412	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	98.820%		
24) Chloroform-d	7.648	84	320225	21.481	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	85.920%		
26) 1,2-Dichloroethane-d4	8.307	65	177563	21.584	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	86.320%		
32) Benzene-d6	8.270	84	575128	26.428	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	105.720%		
36) 1,2-Dichloropropane-d6	9.270	67	193190	28.499	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	114.000%		
41) Toluene-d8	10.325	98	413860	22.154	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	88.600%		
43) trans-1,3-Dichloroprop...	10.575	79	41249	14.738	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	58.960%		
47) 2-Hexanone-d5	10.922	63	45288	63.581	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	127.160%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	151301	28.263	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	113.040%		
66) 1,2-Dichlorobenzene-d4	13.848	152	82825	26.232	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	104.920%		
Target Compounds					Qvalue	
13) Acetone	4.131	43	155367	74.460	ug/L	99
14) Carbon disulfide	4.387	76	143356	6.070	ug/L	100
20) cis-1,2-Dichloroethene	7.173	96	7654	0.972	ug/L	90
22) 2-Butanone	7.179	43	41253	13.416	ug/L	95
33) Benzene	8.319	78	29211	1.270	ug/L	100
34) Trichloroethene	9.093	95	7425	1.268	ug/L	93
42) Toluene	10.386	91	16526	0.710	ug/L	98
52) Ethylbenzene	11.733	91	14524	0.557	ug/L	100
53) m,p-Xylene	11.831	106	13065	1.352	ug/L	89
54) o-Xylene	12.160	106	6071	0.673	ug/L	91
63) 1,2,4-Trimethylbenzene	13.251	105	10386	1.022	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082323\
Data File : VW026841.D
Acq On : 24 Aug 2023 01:45
Operator : SY/MD
Sample : 04113-03
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EZYH6

Quant Time: Aug 24 04:50:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081023SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Aug 23 01:29:10 2023
Response via : Initial Calibration









#13
Acetone
Concen: 74.460 ug/L
RT: 4.131 min Scan# 301
Delta R.T. 0.006 min
Lab File: VW026841.D
Acq: 24 Aug 2023 01:45




Instrument :
MSVOA_W
ClientSampleId :
EZYH6


Tgt Ion: 43 Resp: 155367
Ion Ratio Lower Upper
43 100
58 29.5 0.0 57.4



#14
Carbon disulfide
Concen: 6.070 ug/L
RT: 4.387 min Scan# 410
Delta R.T. -0.012 min
Lab File: VW026841.D
Acq: 24 Aug 2023 01:45

Tgt Ion: 76 Resp: 143356
Ion Ratio Lower Upper
76 100
78 9.0 7.1 10.7









#20
 cis-1,2-Dichloroethene
 Concen: 0.972 ug/L
 RT: 7.173 min Scan# 866
 Delta R.T. 0.006 min
 Lab File: VW026841.D
 Acq: 24 Aug 2023 01:45




Instrument :
 MSVOA_W
 ClientSampleId :
 EZYH6


Tgt Ion: 96 Resp: 7654
 Ion Ratio Lower Upper
 96 100
 61 128.7 101.2 188.0
 98 59.9 43.8 81.3



#22
 2-Butanone
 Concen: 13.416 ug/L
 RT: 7.179 min Scan# 868
 Delta R.T. 0.006 min
 Lab File: VW026841.D
 Acq: 24 Aug 2023 01:45

Tgt Ion: 43 Resp: 41253
 Ion Ratio Lower Upper
 43 100
 72 25.9 11.8 35.4





#33

Benzene

Concen: 1.270 ug/L

RT: 8.319 min Scan# 1055

Delta R.T. -0.006 min

Lab File: VW026841.D

Acq: 24 Aug 2023 01:45





Instrument :

MSVOA_W

ClientSampleId :

EZYH6

Tgt Ion: 78 Resp: 29211



#34

Trichloroethene

Concen: 1.268 ug/L

RT: 9.093 min Scan# 1182

Delta R.T. -0.000 min

Lab File: VW026841.D

Acq: 24 Aug 2023 01:45

Tgt Ion: 95 Resp: 7425




Ion Ratio Lower Upper


95 100

97 67.0 44.6 82.8

132 88.9 66.4 123.2

130 86.3 67.3 125.1





#42

Toluene

Concen: 0.710 ug/L

RT: 10.386 min Scan# 1

Delta R.T. -0.000 min

Lab File: VW026841.D

Acq: 24 Aug 2023 01:45

Instrument :

MSVOA_W

ClientSampleId :





EZYH6

Tgt Ion: 91 Resp: 16526

Ion Ratio Lower Upper

91 100

92 60.0 40.8 75.8



#52

Ethylbenzene

Concen: 0.557 ug/L

RT: 11.733 min Scan# 1615

Delta R.T. 0.006 min

Lab File: VW026841.D




Acq: 24 Aug 2023 01:45


Tgt Ion: 91 Resp: 14524

Ion Ratio Lower Upper

91 100

106 28.6 20.2 37.6









#53
m,p-Xylene
Concen: 1.352 ug/L
RT: 11.831 min Scan# 1
Delta R.T. -0.006 min
Lab File: VW026841.D
Acq: 24 Aug 2023 01:45




Instrument :
MSVOA_W
ClientSampleId :
EZYH6


Tgt Ion:106 Resp: 13065
Ion Ratio Lower Upper
106 100
91 201.6 152.9 284.1



#54
o-Xylene
Concen: 0.673 ug/L
RT: 12.160 min Scan# 1685
Delta R.T. -0.006 min
Lab File: VW026841.D
Acq: 24 Aug 2023 01:45

Tgt Ion:106 Resp: 6071
Ion Ratio Lower Upper
106 100
91 199.8 149.9 278.3





#63
 1,2,4-Trimethylbenzene
 Concen: 1.022 ug/L
 RT: 13.251 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VW026841.D
 Acq: 24 Aug 2023 01:45

Instrument :
 MSVOA_W
 ClientSampleId :
 EZYH6

Tgt Ion:105 Resp: 10386
 Ion Ratio Lower Upper
 105 100
 120 43.7 36.2 54.4

