

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082621\
 Data File : VW019910.D
 Acq On : 26 Aug 2021 12:23
 Operator : SY/VA
 Sample : VIBLK414
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK414

Quant Time: Aug 27 01:02:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 27 00:59:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	374317	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	363001	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	171231	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	123016	21.211	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.840%
7) Chloroethane-d5	2.898	69	106998	21.019	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.080%
11) 1,1-Dichloroethene-d2	4.025	63	162761	16.173	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.680%
21) 2-Butanone-d5	7.080	46	67500	57.299	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	114.600%
24) Chloroform-d	7.653	84	237142	23.036	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.160%
26) 1,2-Dichloroethane-d4	8.311	65	134981	24.332	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.320%
32) Benzene-d6	8.281	84	482576	23.316	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.280%
36) 1,2-Dichloropropane-d6	9.274	67	146098	23.309	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.240%
41) Toluene-d8	10.323	98	433908	22.877	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.520%
43) trans-1,3-Dichloroprop...	10.579	79	61440	23.728	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.920%
47) 2-Hexanone-d5	10.927	63	47223	54.083	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.160%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	126869	24.870	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	99.480%
66) 1,2-Dichlorobenzene-d4	13.853	152	149927	25.526	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.120%
Target Compounds						Qvalue
16) Methylene chloride	4.922	84	12135	1.889	ug/L	90
42) Toluene	10.390	91	7622	0.323	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082621\
Data File : VW019910.D
Acq On : 26 Aug 2021 12:23
Operator : SY/VA
Sample : VIBLK414
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK414

Quant Time: Aug 27 01:02:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Aug 27 00:59:57 2021
Response via : Initial Calibration

