

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082621\
 Data File : VW019912.D
 Acq On : 26 Aug 2021 13:11
 Operator : SY/VA
 Sample : VIBLK415
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK415

Quant Time: Aug 27 01:03:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 27 00:59:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	375970	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	344236	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	166446	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	108193	18.573	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.280%
7) Chloroethane-d5	2.898	69	97606	19.090	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.360%
11) 1,1-Dichloroethene-d2	4.025	63	141286	13.977	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.920%
21) 2-Butanone-d5	7.092	46	51829	43.803	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.600%
24) Chloroform-d	7.653	84	205807	19.904	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.600%
26) 1,2-Dichloroethane-d4	8.311	65	114559	20.560	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	82.240%
32) Benzene-d6	8.281	84	412734	21.028	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.120%
36) 1,2-Dichloropropane-d6	9.274	67	127941	21.525	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	86.120%
41) Toluene-d8	10.323	98	372543	20.712	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	82.840%
43) trans-1,3-Dichloroprop...	10.579	79	51303	20.893	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.560%
47) 2-Hexanone-d5	10.927	63	36488	44.067	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.140%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	106200	21.953	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.800%
66) 1,2-Dichlorobenzene-d4	13.853	152	127974	22.414	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.640%
Target Compounds						Qvalue
16) Methylene chloride	4.922	84	13495	2.091	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082621\
Data File : VW019912.D
Acq On : 26 Aug 2021 13:11
Operator : SY/VA
Sample : VIBLK415
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK415

Quant Time: Aug 27 01:03:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM081821SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Aug 27 00:59:57 2021
Response via : Initial Calibration

