

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082622\
 Data File : VW024417.D
 Acq On : 27 Aug 2022 21:52
 Operator : SY/VA
 Sample : N4363-14
 Misc : 5.43g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBVW8

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 08/30/2022
 Supervised By :Mahesh Dadoda 08/30/2022

Supervised By :Mahesh
 Dadoda
 08/30/2022

Quant Time: Aug 29 01:51:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Aug 29 01:46:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	171426	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	153111	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	66039	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.337	65	139782	29.580	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	118.320%		
7) Chloroethane-d5	2.867	69	92699	26.182	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	104.720%		
11) 1,1-Dichloroethene-d2	4.007	63	63250m	15.433	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	61.720%		
21) 2-Butanone-d5	7.074	46	32713m	37.585	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	75.160%		
24) Chloroform-d	7.647	84	94348	17.119	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	68.480%		
26) 1,2-Dichloroethane-d4	8.299	65	57782m	17.289	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	69.160%#		
32) Benzene-d6	8.268	84	187069	21.324	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	85.280%		
36) 1,2-Dichloropropane-d6	9.274	67	54535	19.989	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	79.960%		
41) Toluene-d8	10.323	98	166853	20.004	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	80.000%		
43) trans-1,3-Dichloroprop...	10.579	79	21627	16.939	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	67.760%		
47) 2-Hexanone-d5	10.920	63	25190	41.260	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	82.520%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	51203	16.577	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	66.320%		
66) 1,2-Dichlorobenzene-d4	13.853	152	49833	20.038	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	80.160%		
Target Compounds						Qvalue
13) Acetone	4.123	43	5729	11.030	ug/L	95
20) cis-1,2-Dichloroethene	7.165	96	24890	9.339	ug/L	90
34) Trichloroethene	9.091	95	19745	7.859	ug/L	94
46) Tetrachloroethene	10.866	164	8290409	4248.692	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082622\
Data File : VW024417.D
Acq On : 27 Aug 2022 21:52
Operator : SY/VA
Sample : N4363-14
Misc : 5.43g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_W
ClientSampleId :
DBVW8


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 08/30/2022
Supervised By :Mahesh Dadoda 08/30/2022

Quant Time: Aug 29 01:51:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Aug 29 01:46:03 2022
Response via : Initial Calibration


Supervised By :Mahesh
Dadoda
08/30/2022





#13
Acetone
Concen: 11.030 ug/L
RT: 4.123 min Scan# 364
Delta R.T. 0.006 min
Lab File: VW024417.D
Acq: 27 Aug 2022 21:52

Instrument :
MSVOA_W
ClientSampleId :
DBVW8






Tgt Ion: 43 Resp: 5729
Ion Ratio Lower Upper
43 100
58 26.4 0.0 58.4

Manual Integrations
APPROVED




Reviewed By :Semsettin Yesilyurt 08/30/2022
Supervised By :Mahesh Dadoda 08/30/2022


Supervised By :Mahesh
Dadoda
08/30/2022



#20
cis-1,2-Dichloroethene
Concen: 9.339 ug/L
RT: 7.165 min Scan# 863
Delta R.T. -0.000 min
Lab File: VW024417.D
Acq: 27 Aug 2022 21:52

Tgt Ion: 96 Resp: 24890
Ion Ratio Lower Upper
96 100
61 117.4 90.2 167.6
68 0.0 0.0 0.0





#34

Trichloroethene

Concen: 7.859 ug/L

RT: 9.091 min Scan# 1178

Delta R.T. 0.006 min

Lab File: VW024417.D


Acq: 27 Aug 2022 21:52

Instrument :

MSVOA_W

ClientSampleId :

DBVW8



Tgt Ion: 95 Resp: 1974

Ion Ratio Lower Upper

95 100

97 61.2 48.4 89.8

132 96.4 68.6 127.4

130 95.8 72.6 134.8

Manual Integrations

APPROVED




Reviewed By :Semsettin Yesilyurt 08/30/2022

Supervised By :Mahesh Dadoda 08/30/2022

Supervised By :Mahesh

Dadoda

08/30/2022



#46

Tetrachloroethene

Concen: 4248.692 ug/L

RT: 10.866 min Scan# 1470

Delta R.T. 0.006 min

Lab File: VW024417.D

Acq: 27 Aug 2022 21:52

Tgt Ion:164 Resp: 8290409

Ion Ratio Lower Upper

164 100

129 92.3 72.9 135.3

131 82.1 63.7 118.3

166 116.4 86.2 160.0

