

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082719\
 Data File : VW012205.D
 Acq On : 27 Aug 2019 11:58
 Operator : SY/VA
 Sample : K4521-01
 Misc : 4.26G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PX-04

Quant Time: Aug 28 05:29:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	241746	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	399935	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	325538	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	145520	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	143328	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
35) Dibromofluoromethane	7.88	113	117073	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
50) Toluene-d8	10.32	98	490647	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
62) 4-Bromofluorobenzene	12.62	95	293964	86.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	172.32%	
Target Compounds						
20) Methylene Chloride	4.91	84	24967	8.990	ug/l	96
80) 1,3,5-Trimethylbenzene	12.90	105	26174	2.873	ug/l	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082719\
Data File : VW012205.D
Acq On : 27 Aug 2019 11:58
Operator : SY/VA
Sample : K4521-01
Misc : 4.26G/5ML/MSVOA W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PX-04

Quant Time: Aug 28 05:29:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

