

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082719\
 Data File : VW012216.D
 Acq On : 27 Aug 2019 16:52
 Operator : SY/VA
 Sample : K4551-17
 Misc : 5.10G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S402-K1-1.0-1.5-082619

Quant Time: Aug 28 06:38:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	247014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	412266	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	348461	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	145424	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	152350	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
35) Dibromofluoromethane	7.88	113	122514	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
50) Toluene-d8	10.32	98	507361	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
62) 4-Bromofluorobenzene	12.62	95	171253	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
Target Compounds						
8) Diethyl Ether	3.68	74	3141	2.285	ug/l	65
16) Acetone	4.14	43	56416	75.161	ug/l	99
20) Methylene Chloride	4.92	84	31445	11.082	ug/l	98
29) Tetrahydrofuran	7.54	42	5881	9.610	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082719\
Data File : VW012216.D
Acq On : 27 Aug 2019 16:52
Operator : SY/VA
Sample : K4551-17
Misc : 5.10G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S402-K1-1.0-1.5-082619

Quant Time: Aug 28 06:38:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

