

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082719\
 Data File : VW012220.D
 Acq On : 27 Aug 2019 18:36
 Operator : SY/VA
 Sample : K4521-13
 Misc : 3.99G/5ML/MSVOA W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PX-16

Quant Time: Aug 28 06:48:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	242553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	407181	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	338723	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	143641	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	147833	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
35) Dibromofluoromethane	7.88	113	121605	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
50) Toluene-d8	10.32	98	496288	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	12.62	95	164563	47.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.74%	
Target Compounds						
20) Methylene Chloride	4.91	84	28323	10.165	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082719\
Data File : VW012220.D
Acq On : 27 Aug 2019 18:36
Operator : SY/VA
Sample : K4521-13
Misc : 3.99G/5ML/MSVOA W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PX-16

Quant Time: Aug 28 06:48:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

