

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082719\
 Data File : VW012230.D
 Acq On : 28 Aug 2019 03:16
 Operator : SY/VA
 Sample : K4551-08
 Misc : 6.86G/5ML/MSVOA W/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S400-K1-1.0-1.5-082619

Quant Time: Aug 28 07:21:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	280576	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	459108	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	355880	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	116898	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	165830	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
35) Dibromofluoromethane	7.88	113	135568	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
50) Toluene-d8	10.32	98	556004	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	12.62	95	160629	41.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.02%	
Target Compounds						
8) Diethyl Ether	3.67	74	3705	2.373	ug/l	84
16) Acetone	4.12	43	49453	58.003	ug/l	97
20) Methylene Chloride	4.91	84	40936	12.701	ug/l	93
29) Tetrahydrofuran	7.53	42	6805	9.790	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082719\
Data File : VW012230.D
Acq On : 28 Aug 2019 03:16
Operator : SY/VA
Sample : K4551-08
Misc : 6.86G/5ML/MSVOA W/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S400-K1-1.0-1.5-082619

Quant Time: Aug 28 07:21:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

