

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082719\
 Data File : VW012239.D
 Acq On : 28 Aug 2019 07:08
 Operator : SY/VA
 Sample : K4472-11
 Misc : 5.05G/5ML/MSVOA W/SOIL
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T2-S3(0)

Quant Time: Aug 28 08:20:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	288182	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	474565	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	378223	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	133512	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	176755	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
35) Dibromofluoromethane	7.88	113	142833	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
50) Toluene-d8	10.32	98	571679	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
62) 4-Bromofluorobenzene	12.62	95	171908	42.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.92%	
Target Compounds						
20) Methylene Chloride	4.91	84	18713	5.653	ug/l	99
29) Tetrahydrofuran	7.52	42	6289	8.809	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082719\
Data File : VW012239.D
Acq On : 28 Aug 2019 07:08
Operator : SY/VA
Sample : K4472-11
Misc : 5.05G/5ML/MSVOA W/SOIL
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T2-S3(0)

Quant Time: Aug 28 08:20:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

