

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082819\
 Data File : VW012251.D
 Acq On : 28 Aug 2019 13:58
 Operator : SY/VA
 Sample : K4521-01
 Misc : 4.39G/5ML/MSVOA W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PX-04

Quant Time: Aug 28 17:53:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	246656	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	407552	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	335850	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	135976	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	137471	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
35) Dibromofluoromethane	7.88	113	117597	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
50) Toluene-d8	10.32	98	492702	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
62) 4-Bromofluorobenzene	12.62	95	157463	45.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.58%	
Target Compounds						
20) Methylene Chloride	4.91	84	17663	6.234	ug/l	94
29) Tetrahydrofuran	7.54	42	2348	3.843	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082819\
Data File : VW012251.D
Acq On : 28 Aug 2019 13:58
Operator : SY/VA
Sample : K4521-01
Misc : 4.39G/5ML/MSVOA W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PX-04

Quant Time: Aug 28 17:53:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

