

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082819\
 Data File : VW012257.D
 Acq On : 28 Aug 2019 16:32
 Operator : SY/VA
 Sample : K4509-09
 Misc : 4.99G/5ML/MSVOA W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-45-7.5-8.0

Quant Time: Aug 28 18:45:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	242165	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	392630	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	303545	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	101602	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	139519	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
35) Dibromofluoromethane	7.88	113	117793	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
50) Toluene-d8	10.32	98	474914	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	12.62	95	132721	39.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.24%	
Target Compounds						
20) Methylene Chloride	4.92	84	24373	8.761	ug/l	92
64) Tetrachloroethene	10.87	164	6380	3.242	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082819\
Data File : VW012257.D
Acq On : 28 Aug 2019 16:32
Operator : SY/VA
Sample : K4509-09
Misc : 4.99G/5ML/MSVOA W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-45-7.5-8.0

Quant Time: Aug 28 18:45:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

