

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082819\
 Data File : VW012261.D
 Acq On : 28 Aug 2019 18:15
 Operator : SY/VA
 Sample : K4510-03
 Misc : 5.04G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-41-11.5-12.0

Quant Time: Aug 29 06:58:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	235204	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	396681	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	328791	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	135742	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	144605	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
35) Dibromofluoromethane	7.88	113	117088	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
50) Toluene-d8	10.32	98	477943	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
62) 4-Bromofluorobenzene	12.62	95	155872	46.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.12%	
Target Compounds						
16) Acetone	4.14	43	10995	15.384	ug/l	Qvalue 100
20) Methylene Chloride	4.92	84	25162	9.313	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082819\
Data File : VW012261.D
Acq On : 28 Aug 2019 18:15
Operator : SY/VA
Sample : K4510-03
Misc : 5.04G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-41-11.5-12.0

Quant Time: Aug 29 06:58:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

