

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW082819\
 Data File : VW012265.D
 Acq On : 28 Aug 2019 19:58
 Operator : SY/VA
 Sample : K4510-11
 Misc : 5.05G/5ML/MSVOA W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-13-6.5-7.0

Quant Time: Aug 29 07:03:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	118503	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	185222	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	153001	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	60092	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	66105	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
35) Dibromofluoromethane	7.88	113	56330	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
50) Toluene-d8	10.32	98	224650	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
62) 4-Bromofluorobenzene	12.62	95	69607	44.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.10%	
Target Compounds						
20) Methylene Chloride	4.91	84	9551	7.016	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW082819\
Data File : VW012265.D
Acq On : 28 Aug 2019 19:58
Operator : SY/VA
Sample : K4510-11
Misc : 5.05G/5ML/MSVOA W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-13-6.5-7.0

Quant Time: Aug 29 07:03:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

