

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW082819\
 Data File : VW012286.D
 Acq On : 29 Aug 2019 06:46
 Operator : SY/VA
 Sample : K4536-06
 Misc : 5.45G/5ML/MSVOA_W/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-1-25-27

Quant Time: Sep 06 03:53:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	246880	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	399973	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	332941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	134043	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	158103	53.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.74%	
35) Dibromofluoromethane	7.88	113	120200	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
50) Toluene-d8	10.32	98	491947	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	12.62	95	159760	46.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.64%	
Target Compounds						
16) Acetone	4.12	43	43909	58.530	ug/l	97
17) Carbon Disulfide	4.37	76	68254	9.120	ug/l	99
20) Methylene Chloride	4.91	84	25949	9.150	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW082819\
Data File : VW012286.D
Acq On : 29 Aug 2019 06:46
Operator : SY/VA
Sample : K4536-06
Misc : 5.45G/5ML/MSVOA_W/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-1-25-27

Quant Time: Sep 06 03:53:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

