

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW083018\
 Data File : VW005017.D
 Acq On : 30 Aug 2018 14:43
 Operator : SY/AP
 Sample : J4671-01RE
 Misc : 5.02G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TL-SPEEDY-DRYRE

Quant Time: Aug 30 15:20:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 30 01:51:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	260298	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	437471	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	328792	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	115641	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	141447	79.44	ug/l	0.00
Spiked Amount			Recovery	=	158.88%	
35) Dibromofluoromethane	7.89	113	132859	61.29	ug/l	0.00
Spiked Amount			Recovery	=	122.58%	
50) Toluene-d8	10.33	98	343857	28.89	ug/l	0.00
Spiked Amount			Recovery	=	57.78%	
62) 4-Bromofluorobenzene	12.62	95	120575	27.79	ug/l	0.00
Spiked Amount			Recovery	=	55.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.15	43	104245	259.62	ug/l	99
20) Methylene Chloride	4.92	84	53121	15.52	ug/l	98
29) Tetrahydrofuran	7.54	42	40749	128.83	ug/l	98
48) Methyl methacrylate	9.45	41	9584	8.46	ug/l	98
51) 4-Methyl-2-Pentanone	10.22	43	38475	19.32	ug/l	98
59) 2-Hexanone	10.98	43	9315	6.46	ug/l	86
84) 1,2,4-Trimethylbenzene	13.26	105	35028	5.07	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW083018\
Data File : VW005017.D
Acq On : 30 Aug 2018 14:43
Operator : SY/AP
Sample : J4671-01RE
Misc : 5.02G/5ML/MSVOA W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TL-SPEEDY-DRYRE

Quant Time: Aug 30 15:20:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082918S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 30 01:51:13 2018
Response via : Initial Calibration

