

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW083018\
 Data File : VW005034.D
 Acq On : 30 Aug 2018 21:55
 Operator : SY/AP
 Sample : J4710-04
 Misc : 5.36G/5ML/MSVOA W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ENY-SB-07-10-10.5

Quant Time: Aug 31 07:01:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 30 01:51:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	389256	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	560468	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	568949	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	269769	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	118265	44.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.84%	
35) Dibromofluoromethane	7.89	113	111962	40.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.62%	
50) Toluene-d8	10.33	98	503727	33.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.06%	
62) 4-Bromofluorobenzene	12.62	95	170664	30.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	61.40%	
Target Compounds						
16) Acetone	4.15	43	36385	60.59	ug/l	98
20) Methylene Chloride	4.92	84	37886	4.76	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW083018\
Data File : VW005034.D
Acq On : 30 Aug 2018 21:55
Operator : SY/AP
Sample : J4710-04
Misc : 5.36G/5ML/MSVOA W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ENY-SB-07-10-10.5

Quant Time: Aug 31 07:01:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082918S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 30 01:51:13 2018
Response via : Initial Calibration

