

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW083019\
 Data File : VW012382.D
 Acq On : 31 Aug 2019 06:22
 Operator : SY/VA
 Sample : K4554-07
 Misc : 4.23G/5ML/MSVOA W/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-15-0-2

Quant Time: Aug 31 06:50:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	206343	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	340835	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	285626	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	118323	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	128897	52.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.10%	
35) Dibromofluoromethane	7.88	113	105477	51.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.70%	
50) Toluene-d8	10.32	98	418848	50.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.54%	
62) 4-Bromofluorobenzene	12.62	95	134193	46.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.30%	
Target Compounds						
20) Methylene Chloride	4.90	84	12821	5.409	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW083019\
Data File : VW012382.D
Acq On : 31 Aug 2019 06:22
Operator : SY/VA
Sample : K4554-07
Misc : 4.23G/5ML/MSVOA W/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-15-(0-2)

Quant Time: Aug 31 06:50:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

