

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW083019\
 Data File : VW012383.D
 Acq On : 31 Aug 2019 06:48
 Operator : SY/VA
 Sample : K4568-09RE
 Misc : 5.08G/5ML/MSVOA W/SOIL
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Aug 31 11:20:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	177320	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	286844	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	182317	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	36779	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	124701	59.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.32%	
35) Dibromofluoromethane	7.88	113	97642	57.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.06%	
50) Toluene-d8	10.32	98	331879	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
62) 4-Bromofluorobenzene	12.62	95	59265	24.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	48.44%	
Target Compounds						
20) Methylene Chloride	4.90	84	10400	5.106	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW083019\
Data File : VW012383.D
Acq On : 31 Aug 2019 06:48
Operator : SY/VA
Sample : K4568-09RE
Misc : 5.08G/5ML/MSVOA W/SOIL
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Aug 31 11:20:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

