

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW090118\
 Data File : VW005067.D
 Acq On : 01 Sep 2018 06:09
 Operator : SY/AP
 Sample : J4270-05RE
 Misc : 3.09G/5ML/MSVOA_W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BU-02-083018

Quant Time: Sep 03 23:53:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	53305	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	81919	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	67851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	22009	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	24174	54.71	ug/l	0.00
Spiked Amount			Recovery	=	109.42%	
35) Dibromofluoromethane	7.95	113	35	0.07	ug/l	0.05
Spiked Amount			Recovery	=	0.14%	
50) Toluene-d8	10.33	98	66152	30.33	ug/l	0.00
Spiked Amount			Recovery	=	60.66%	
62) 4-Bromofluorobenzene	12.62	95	17979	23.37	ug/l	0.00
Spiked Amount			Recovery	=	46.74%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.78	94	954	Below Cal	#	58
10) Methyl Iodide	4.29	142	1864	6.278	ug/l	97
11) Tert butyl alcohol	5.15	59	239	Below Cal	#	72
16) Acetone	4.13	43	9011	101.685	ug/l	94
25) 2-Butanone	7.18	43	1283	8.347	ug/l	95
31) Cyclohexane	7.97	56	937	1.080	ug/l #	16
36) 1,1-Dichloropropene	7.96	75	3596	4.609	ug/l #	47
38) Carbon Tetrachloride	7.96	117	4624	6.233	ug/l #	19
48) Methyl methacrylate	9.42	41	400	1.565	ug/l #	23
59) 2-Hexanone	10.96	43	387	1.608	ug/l	77
75) 1,1,2,2-Tetrachloroethane	12.71	83	255	1.079	ug/l #	35
76) 1,2,3-Trichloropropane	12.77	75	371	2.246	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.62	75	7947	115.781	ug/l #	14
84) 1,2,4-Trimethylbenzene	13.26	105	1445	1.057	ug/l	92
87) 1,3-Dichlorobenzene	13.50	146	1444	1.838	ug/l	94
88) 1,4-Dichlorobenzene	13.50	146	1444	1.853	ug/l	94
89) n-Butylbenzene	13.84	91	1851	1.370	ug/l	95
90) Hexachloroethane	14.10	117	280	1.156	ug/l	88
91) 1,2-Dichlorobenzene	13.88	146	1173	1.713	ug/l	87
92) 1,2-Dibromo-3-Chloropropan	14.49	75	75	2.032	ug/l #	67
93) 1,2,4-Trichlorobenzene	15.14	180	1542	3.359	ug/l	96
94) Hexachlorobutadiene	15.24	225	726	2.543	ug/l	93
95) Naphthalene	15.38	128	3263	7.764	ug/l	98
96) 1,2,3-Trichlorobenzene	15.57	180	1377	3.505	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW090118\
Data File : VW005067.D
Acq On : 01 Sep 2018 06:09
Operator : SY/AP
Sample : J4270-05RE
Misc : 3.09G/5ML/MSVOA_W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BU-02-083018

Quant Time: Sep 03 23:53:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.96 min Scan# 1045

Delta R.T. -0.00 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09

Instrument :

MSVOA_W

ClientSampled :


BU-02-083018

Tgt Ion: 168 Resp: 53305

Ion Ratio Lower Upper




168 100

99 46.8 39.4 59.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW



#5

Bromomethane

Concen: Below Cal

RT: 2.78 min Scan# 196

Delta R.T. 0.01 min

Lab File: VW005067.D


Acq: 01 Sep 2018 06:09

Tgt Ion: 94 Resp: 954

Ion Ratio Lower Upper



94 100


96 51.3 73.2 109.8#



Abundance Ion 93.90 (93.60 to 94.60): VW

Ion 95.90 (95.60 to 96.60): VW









#10
Methyl Iodide
Concen: 6.278 ug/l
RT: 4.29 min Scan# 443
Delta R.T. 0.01 min
Lab File: VW005067.D
Acq: 01 Sep 2018 06:09

Instrument :
MSVOA_W
ClientSampleId :
BU-02-083018

Tgt Ion:142 Resp: 1864
Ion Ratio Lower Upper
142 100
127 35.8 30.8 46.2
141 14.2 11.6 17.4




Abundance Ion 141.80 (141.50 to 142.50): V
Ion 126.80 (126.50 to 127.50): V
Ion 140.80 (140.50 to 141.50): V






#16
Acetone
Concen: 101.685 ug/l
RT: 4.13 min Scan# 417
Delta R.T. -0.01 min
Lab File: VW005067.D
Acq: 01 Sep 2018 06:09

Tgt Ion: 43 Resp: 9011
Ion Ratio Lower Upper
43 100
58 40.0 29.0 43.6



Abundance Ion 43.00 (42.70 to 43.70): VW
Ion 58.00 (57.70 to 58.70): VW





#25

2-Butanone

Concen: 8.347 ug/l

RT: 7.18 min Scan# 918

Delta R.T. -0.00 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09

Instrument :

MSVOA_W

ClientSampled :


BU-02-083018

Tgt Ion: 43 Resp: 1283

Ion Ratio Lower Upper




43 100

72 26.4 23.1 34.7



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 72.00 (71.70 to 72.70): VW



#31

Cyclohexane

Concen: 1.080 ug/l

RT: 7.97 min Scan# 1047

Delta R.T. 0.01 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09


Tgt Ion: 56 Resp: 937

Ion Ratio Lower Upper

56 100

69 166.6 28.2 42.2#



84 116.4 71.8 107.8#




Abundance Ion 56.00 (55.70 to 56.70): VW

Ion 69.00 (68.70 to 69.70): VW

Ion 84.00 (83.70 to 84.70): VW









#33
 1,2-Dichloroethane-d4
 Concen: 54.708 ug/l
 RT: 8.31 min Scan# 1103
 Delta R.T. 0.00 min
 Lab File: VW005067.D
 Acq: 01 Sep 2018 06:09




Instrument :
 MSVOA_W
 ClientSampleId :
 BU-02-083018


Tgt Ion: 65 Resp: 24174
 Ion Ratio Lower Upper
 65 100
 67 54.0 0.0 109.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.85 min Scan# 1191
 Delta R.T. 0.00 min
 Lab File: VW005067.D
 Acq: 01 Sep 2018 06:09

Tgt Ion: 114 Resp: 81919
 Ion Ratio Lower Upper
 114 100
 63 18.6 0.0 38.0
 88 14.7 0.0 30.2









#35
 Dibromofluoromethane
 Concen: 0.074 ug/l
 RT: 7.95 min Scan# 1043
 Delta R.T. 0.05 min
 Lab File: VW005067.D
 Acq: 01 Sep 2018 06:09

Instrument :
 MSVOA_W
 ClientSampleId :
 BU-02-083018

Tgt Ion: 113 Resp: 35
 Ion Ratio Lower Upper
 113 100
 111 0.0 82.2 123.2#
 192 0.0 18.1 27.1#




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#36
 1,1-Dichloropropene
 Concen: 4.609 ug/l
 RT: 7.96 min Scan# 1045
 Delta R.T. -0.13 min
 Lab File: VW005067.D
 Acq: 01 Sep 2018 06:09

Tgt Ion: 75 Resp: 3596
 Ion Ratio Lower Upper
 75 100
 110 7.9 18.9 56.7#
 77 0.0 25.5 38.3#



Abundance Ion 74.90 (74.60 to 75.60): VW
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): VW





#38

Carbon Tetrachloride

Concen: 6.233 ug/l

RT: 7.96 min Scan# 1045

Delta R.T. -0.12 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09

Instrument :

MSVOA_W

ClientSampleId :

BU-02-083018


Tgt Ion: 117 Resp: 4624

Ion Ratio Lower Upper

117 100

119 9.3 76.6 115.0#




121 0.0 24.8 37.2#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#48

Methyl methacrylate

Concen: 1.565 ug/l

RT: 9.42 min Scan# 1284

Delta R.T. -0.03 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09


Tgt Ion: 41 Resp: 400

Ion Ratio Lower Upper

41 100

69 0.0 80.4 120.6#



39 29.8 43.5 65.3#




Abundance Ion 41.00 (40.70 to 41.70): VW

Ion 69.00 (68.70 to 69.70): VW

Ion 39.00 (38.70 to 39.70): VW





#50

Toluene-d8

Concen: 30.327 ug/l

RT: 10.33 min Scan# 1434

Delta R.T. 0.00 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09

Instrument :

MSVOA_W

ClientSampleId :





BU-02-083018

Tgt Ion: 98 Resp: 66152

Ion Ratio Lower Upper

98 100

100 64.0 51.7 77.5



#59

2-Hexanone

Concen: 1.608 ug/l

RT: 10.96 min Scan# 1538

Delta R.T. -0.01 min

Lab File: VW005067.D




Acq: 01 Sep 2018 06:09


Tgt Ion: 43 Resp: 387

Ion Ratio Lower Upper

43 100

58 40.8 28.7 86.3





#62

4-Bromofluorobenzene

Concen: 23.375 ug/l

RT: 12.62 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09

Instrument :

MSVOA_W

ClientSampleId :

BU-02-083018


Tgt Ion: 95 Resp: 17979

Ion Ratio Lower Upper

95 100

174 89.5 0.0 173.4




176 84.1 0.0 167.4



Abundance Ion 94.90 (94.60 to 95.60): VW

Ion 173.80 (173.50 to 174.50): VW

Ion 175.80 (175.50 to 176.50): VW



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.63 min Scan# 1648

Delta R.T. -0.01 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09


Tgt Ion: 117 Resp: 67851

Ion Ratio Lower Upper

117 100

82 53.1 40.9 61.3



119 29.6 24.9 37.3




Abundance Ion 116.90 (116.60 to 117.60): VW

Ion 82.00 (81.70 to 82.70): VW

Ion 118.90 (118.60 to 119.60): VW









#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.57 min Scan# 1965
 Delta R.T. 0.00 min
 Lab File: VW005067.D
 Acq: 01 Sep 2018 06:09

Instrument :
 MSVOA_W
 ClientSampleId :
 BU-02-083018

Tgt Ion: 152 Resp: 22009
 Ion Ratio Lower Upper
 152 100
 115 53.7 27.8 83.4
 150 158.1 0.0 351.4




Abundance Ion 151.90 (151.60 to 152.60): V
 Ion 114.90 (114.60 to 115.60): V
 Ion 149.90 (149.60 to 150.60): V






#75
 1,1,2,2-Tetrachloroethane
 Concen: 1.079 ug/l
 RT: 12.71 min Scan# 1825
 Delta R.T. -0.01 min
 Lab File: VW005067.D
 Acq: 01 Sep 2018 06:09

Tgt Ion: 83 Resp: 255
 Ion Ratio Lower Upper
 83 100
 131 0.0 5.5 16.4#
 85 8.6 32.5 97.4#



Abundance Ion 82.90 (82.60 to 83.60): VW
 Ion 130.80 (130.50 to 131.50): V
 Ion 84.90 (84.60 to 85.60): VW





#76

1,2,3-Trichloropropane

Concen: 2.246 ug/l

RT: 12.77 min Scan# 1834

Delta R.T. -0.01 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09

Instrument :

MSVOA_W

ClientSampleId :


BU-02-083018

Tgt Ion: 75 Resp: 371

Ion Ratio Lower Upper




75 100

77 0.0 0.0 0.0



Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 77.00 (76.70 to 77.70): VW



#81

trans-1,4-Dichloro-2-butene

Concen: 115.781 ug/l

RT: 12.62 min Scan# 1810

Delta R.T. 0.10 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09


Tgt Ion: 75 Resp: 7947

Ion Ratio Lower Upper

75 100

53 0.0 69.8 104.8#



89 0.0 36.0 54.0#




Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 53.00 (52.70 to 53.70): VW

Ion 88.90 (88.60 to 89.60): VW









#84
 1,2,4-Trimethylbenzene
 Concen: 1.057 ug/l
 RT: 13.26 min Scan# 1914
 Delta R.T. 0.00 min
 Lab File: VW005067.D
 Acq: 01 Sep 2018 06:09




Instrument :
 MSVOA_W
 ClientSampleId :
 BU-02-083018


Tgt Ion:105 Resp: 1445
 Ion Ratio Lower Upper
 105 100
 120 41.1 23.1 69.3



#87
 1,3-Dichlorobenzene
 Concen: 1.838 ug/l
 RT: 13.50 min Scan# 1954
 Delta R.T. -0.01 min
 Lab File: VW005067.D
 Acq: 01 Sep 2018 06:09

Tgt Ion:146 Resp: 1444
 Ion Ratio Lower Upper
 146 100
 111 43.8 19.2 57.6
 148 66.6 31.9 95.7





#88

1,4-Dichlorobenzene

Concen: 1.853 ug/l

RT: 13.50 min Scan# 1954

Delta R.T. -0.09 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09

Instrument :

MSVOA_W

ClientSampleId :

BU-02-083018


Tgt Ion:146 Resp: 1444

Ion Ratio Lower Upper

146 100

111 43.8 18.6 56.0




148 66.6 32.0 96.0



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#89

n-Butylbenzene

Concen: 1.370 ug/l

RT: 13.84 min Scan# 2009

Delta R.T. 0.01 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09


Tgt Ion: 91 Resp: 1851

Ion Ratio Lower Upper

91 100

92 49.7 26.5 79.5



134 29.4 13.5 40.4




Abundance Ion 91.00 (90.70 to 91.70): VW

Ion 92.00 (91.70 to 92.70): VW

Ion 134.00 (133.70 to 134.70): V





#90

Hexachloroethane

Concen: 1.156 ug/l

RT: 14.10 min Scan# 2052

Delta R.T. 0.00 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09

Instrument :

MSVOA_W

ClientSampleId :


BU-02-083018

Tgt Ion:117 Resp: 280

Ion Ratio Lower Upper




117 100

201 77.9 44.6 133.8



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



#91

1,2-Dichlorobenzene

Concen: 1.713 ug/l

RT: 13.88 min Scan# 2017

Delta R.T. 0.01 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09


Tgt Ion:146 Resp: 1173

Ion Ratio Lower Upper

146 100

111 48.7 19.9 59.7



148 74.0 31.9 95.9




Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V





#92

1,2-Dibromo-3-Chloropropane

Concen: 2.032 ug/l

RT: 14.49 min Scan# 2117

Delta R.T. 0.00 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09

Instrument :

MSVOA_W

ClientSampled :

BU-02-083018


Tgt Ion: 75 Resp: 75

Ion Ratio Lower Upper

75 100

155 41.3 46.0 137.8#




157 98.7 58.6 175.7



Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V



#93

1,2,4-Trichlorobenzene

Concen: 3.359 ug/l

RT: 15.14 min Scan# 2223

Delta R.T. 0.00 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09


Tgt Ion: 180 Resp: 1542

Ion Ratio Lower Upper

180 100

182 98.2 47.5 142.6



145 25.6 14.4 43.2




Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.80 (144.50 to 145.50): V





#94

Hexachlorobutadiene

Concen: 2.543 ug/l

RT: 15.24 min Scan# 2240

Delta R.T. 0.00 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09

Instrument :

MSVOA_W

ClientSampleId :

BU-02-083018


Tgt Ion: 225 Resp: 726

Ion Ratio Lower Upper




225 100

223 62.4 31.4 94.3

227 75.1 32.0 96.0



Abundance Ion 224.70 (224.40 to 225.40): V
Ion 222.70 (222.40 to 223.40): V
Ion 226.70 (226.40 to 227.40): V



#95

Naphthalene

Concen: 7.764 ug/l

RT: 15.38 min Scan# 2262

Delta R.T. 0.00 min

Lab File: VW005067.D

Acq: 01 Sep 2018 06:09


Tgt Ion: 128 Resp: 3263

Ion Ratio Lower Upper



128 100


127 13.0 10.4 15.6

129 9.2 8.7 13.1



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V





#96
 1,2,3-Trichlorobenzene
 Concen: 3.505 ug/l
 RT: 15.57 min Scan# 2293
 Delta R.T. 0.00 min
 Lab File: VW005067.D
 Acq: 01 Sep 2018 06:09

Instrument :
 MSVOA_W
 ClientSampleId :
 BU-02-083018

Tgt Ion	Ratio	Lower	Upper
180	100		
182	95.7	47.4	142.3
145	37.5	15.4	46.1

