

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090118\
 Data File : VW005066.D
 Acq On : 01 Sep 2018 05:43
 Operator : SY/AP
 Sample : J4275-21
 Misc : 5.39G/5ML/MSVOA W/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CL-02-082918-B

Quant Time: Sep 04 00:24:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	441233	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	692830	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	617923	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	255369	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	157637	43.10	ug/l	0.00
Spiked Amount			Recovery	=	86.20%	
35) Dibromofluoromethane	7.89	113	157293	39.29	ug/l	0.00
Spiked Amount			Recovery	=	78.58%	
50) Toluene-d8	10.33	98	597265	32.37	ug/l	0.00
Spiked Amount			Recovery	=	64.74%	
62) 4-Bromofluorobenzene	12.62	95	189196	29.08	ug/l	0.00
Spiked Amount			Recovery	=	58.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.13	43	69928	95.33	ug/l	94
17) Carbon Disulfide	4.38	76	34636	2.79	ug/l	98
25) 2-Butanone	7.18	43	10318	8.11	ug/l	98
40) Benzene	8.33	78	30009	1.49	ug/l	95
68) m/p-Xylenes	11.84	106	27223	2.89	ug/l	99
69) o-Xylene	12.17	106	15220	1.75	ug/l	95
84) 1,2,4-Trimethylbenzene	13.26	105	26610	1.68	ug/l	99
95) Naphthalene	15.38	128	99141	13.89	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090118\
Data File : VW005066.D
Acq On : 01 Sep 2018 05:43
Operator : SY/AP
Sample : J4275-21
Misc : 5.39G/5ML/MSVOA W/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CL-02-082918-B

Quant Time: Sep 04 00:24:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

