

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090118\
 Data File : VW005068.D
 Acq On : 01 Sep 2018 06:34
 Operator : SY/AP
 Sample : J4730-02RE
 Misc : 8.22G/5ML/MSVOA W/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FK-BPM-02-008-ABCDRE

Quant Time: Sep 04 00:40:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	421340	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	669440	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	647717	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	310809	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	166771	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
35) Dibromofluoromethane	7.89	113	165147	42.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.38%	
50) Toluene-d8	10.33	98	616883	34.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.22%	
62) 4-Bromofluorobenzene	12.63	95	218196	34.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.42%	
Target Compounds						
16) Acetone	4.13	43	8994	12.84	ug/l	97
95) Naphthalene	15.38	128	7410	4.59	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090118\
Data File : VW005068.D
Acq On : 01 Sep 2018 06:34
Operator : SY/AP
Sample : J4730-02RE
Misc : 8.22G/5ML/MSVOA W/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FK-BPM-02-008-ABCDRE

Quant Time: Sep 04 00:40:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

