

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090118\
 Data File : VW005071.D
 Acq On : 01 Sep 2018 07:50
 Operator : SY/AP
 Sample : J4654-03RE
 Misc : 2.00G/5ML/MSVOA W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ENY-SB-02-1.5-2RE

Quant Time: Sep 04 00:59:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	297431	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	386290	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	280267	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	77051	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	118684	48.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.28%	
35) Dibromofluoromethane	7.89	113	123569	55.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.70%	
50) Toluene-d8	10.33	98	322561	31.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	62.72%	
62) 4-Bromofluorobenzene	12.62	95	56953	15.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	31.40%	
Target Compounds						
16) Acetone	4.14	43	33250	67.24	ug/l	96
20) Methylene Chloride	4.92	84	30559	5.70	ug/l	98
95) Naphthalene	15.38	128	11785	7.88	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090118\
Data File : VW005071.D
Acq On : 01 Sep 2018 07:50
Operator : SY/AP
Sample : J4654-03RE
Misc : 2.00G/5ML/MSVOA W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ENY-SB-02-1.5-2RE

Quant Time: Sep 04 00:59:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

