

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090118\
 Data File : VW005072.D
 Acq On : 01 Sep 2018 08:16
 Operator : SY/AP
 Sample : J4654-04RE
 Misc : 3.05G/5ML/MSVOA W/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ENY-SB-04-1.5-2RE

Quant Time: Sep 04 01:03:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	223818	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	280556	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	154109	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	18492	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	132305	71.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	142.62%	
35) Dibromofluoromethane	7.89	113	134215	82.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	165.56%	
50) Toluene-d8	10.33	98	254384	34.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.10%	
62) 4-Bromofluorobenzene	12.62	95	22408	8.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	17.02%	
Target Compounds						
16) Acetone	4.14	43	22461	60.36	ug/l	96
20) Methylene Chloride	4.92	84	28319	8.00	ug/l	94
95) Naphthalene	15.38	128	2755	7.78	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090118\
Data File : VW005072.D
Acq On : 01 Sep 2018 08:16
Operator : SY/AP
Sample : J4654-04RE
Misc : 3.05G/5ML/MSVOA W/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ENY-SB-04-1.5-2RE

Quant Time: Sep 04 01:03:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

