

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090119\
 Data File : VW012443.D
 Acq On : 01 Sep 2019 13:05
 Operator : SY/VA
 Sample : K4475-11RE
 Misc : 4.89G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T1-2-B-(0)RE

Quant Time: Sep 02 07:41:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	185326	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	301872	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	213279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	54911	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	126401	57.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.74%	
35) Dibromofluoromethane	7.88	113	95996	53.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.56%	
50) Toluene-d8	10.32	98	353662	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
62) 4-Bromofluorobenzene	12.62	95	82530	32.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.10%	
Target Compounds						
20) Methylene Chloride	4.91	84	12897	6.058	ug/l	96
29) Tetrahydrofuran	7.54	42	4264	9.287	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090119\
Data File : VW012443.D
Acq On : 01 Sep 2019 13:05
Operator : SY/VA
Sample : K4475-11RE
Misc : 4.89G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T1-2-B-(0)RE

Quant Time: Sep 02 07:41:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

