

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090119\
 Data File : VW012447.D
 Acq On : 01 Sep 2019 14:48
 Operator : SY/VA
 Sample : K4517-09
 Misc : 4.85G/5ML/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T4-S-D2-(2.5)

Quant Time: Sep 02 07:49:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	193990	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	323141	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	272954	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	112567	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	131210	56.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.80%	
35) Dibromofluoromethane	7.88	113	100475	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
50) Toluene-d8	10.32	98	397274	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
62) 4-Bromofluorobenzene	12.62	95	125139	45.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.78%	
Target Compounds						
16) Acetone	4.12	43	29327	49.751	ug/l	95
20) Methylene Chloride	4.92	84	27535	12.356	ug/l	92
29) Tetrahydrofuran	7.53	42	4326	9.001	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090119\
Data File : VW012447.D
Acq On : 01 Sep 2019 14:48
Operator : SY/VA
Sample : K4517-09
Misc : 4.85G/5ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T4-S-D2-(2.5)

Quant Time: Sep 02 07:49:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

