

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090119\
 Data File : VW012478.D
 Acq On : 02 Sep 2019 05:24
 Operator : SY/VA
 Sample : K4528-04
 Misc : 3.00G/5ML/MSVOA W/SOIL
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T9-12-13-Q2-(1.9)

Quant Time: Sep 02 08:46:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	233711	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	383243	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	309401	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	110050	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	153096	55.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.22%	
35) Dibromofluoromethane	7.88	113	119788	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
50) Toluene-d8	10.32	98	464434	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
62) 4-Bromofluorobenzene	12.62	95	141587	43.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.60%	
Target Compounds						
20) Methylene Chloride	4.90	84	39963	14.885	ug/l	95
29) Tetrahydrofuran	7.52	42	5729	9.895	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090119\
Data File : VW012478.D
Acq On : 02 Sep 2019 05:24
Operator : SY/VA
Sample : K4528-04
Misc : 3.00G/5ML/MSVOA W/SOIL
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T9-12-13-Q2-(1.9)

Quant Time: Sep 02 08:46:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

