

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090119\
 Data File : VW012483.D
 Acq On : 02 Sep 2019 07:33
 Operator : SY/VA
 Sample : K4528-11
 Misc : 5.54G/5ML/MSVOA W/SOIL
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T9-12-13-S1-(0)

Quant Time: Sep 02 08:52:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	196484	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	318320	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	259525	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	88110	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	129424	55.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.82%	
35) Dibromofluoromethane	7.88	113	100598	52.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.88%	
50) Toluene-d8	10.32	98	385652	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
62) 4-Bromofluorobenzene	12.62	95	114006	41.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.96%	
Target Compounds						
20) Methylene Chloride	4.92	84	16771	7.430	ug/l	97
29) Tetrahydrofuran	7.54	42	4090	8.402	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090119\
Data File : VW012483.D
Acq On : 02 Sep 2019 07:33
Operator : SY/VA
Sample : K4528-11
Misc : 5.54G/5ML/MSVOA W/SOIL
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T9-12-13-S1-(0)

Quant Time: Sep 02 08:52:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

