


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW090119\
 Data File : VW012437.D
 Acq On : 01 Sep 2019 10:00
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Title : SW846 8260
 Last Update : Mon Aug 26 07:18:31 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	27256	PASS
75	95	30	60	53.8	65125	PASS
95	95	100	100	100.0	120973	PASS
96	95	5	9	6.8	8217	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	78.7	95216	PASS
175	174	5	9	8.1	7673	PASS
176	174	95	101	97.2	92546	PASS
177	176	5	9	6.8	6288	PASS