



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_W\Data\VW090123\  
Data File : VW026982.D  
Acq On : 01 Sep 2023 08:26  
Operator : SY/MD  
Sample : BFB575  
Misc : 5.00g/10mL/MSVOA\_W/SOIL  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_W  
ClientSampleId :  
BFB575

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_W\Method\SFAMWLM081023SMA.M  
Title : SFAM01.0  
Last Update : Sat Sep 02 00:33:01 2023



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1751

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw    | Result    |
|--------|---------|--------|--------|-------|--------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn    | Pass/Fail |
| 50     | 95      | 15     | 40     | 17.7  | 61809  | PASS      |
| 75     | 95      | 30     | 80     | 49.6  | 173013 | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 349126 | PASS      |
| 96     | 95      | 5      | 9      | 6.7   | 23419  | PASS      |
| 173    | 174     | 0.00   | 2      | 0.0   | 0      | PASS      |
| 174    | 95      | 50     | 120    | 74.6  | 260416 | PASS      |
| 175    | 174     | 5      | 9      | 5.3   | 13925  | PASS      |
| 176    | 174     | 95     | 105    | 98.3  | 255872 | PASS      |
| 177    | 176     | 5      | 10     | 6.9   | 17645  | PASS      |

| m/z    | Abundance |
|--------|-----------|
| 35.90  | 473.0     |
| 37.40  | 140.0     |
| 38.00  | 295.0     |
| 38.60  | 150.0     |
| 39.30  | 182.0     |
| 40.00  | 639.0     |
| 40.90  | 143.0     |
| 41.30  | 175.0     |
| 42.10  | 209.0     |
| 43.20  | 283.0     |
| 44.00  | 802.0     |
| 44.80  | 102.0     |
| 45.30  | 138.0     |
| 46.40  | 63.0      |
| 47.30  | 83.0      |
| 48.40  | 124.0     |
| 49.50  | 93.0      |
| 50.00  | 106.0     |
| 50.60  | 140.0     |
| 52.00  | 76.0      |
| 52.50  | 54.0      |
| 53.30  | 124.0     |
| 54.10  | 89.0      |
| 55.10  | 251.0     |
| 55.90  | 175.0     |
| 57.00  | 263.0     |
| 57.80  | 65.0      |
| 58.90  | 62.0      |
| 59.60  | 55.0      |
| 60.40  | 78.0      |
| 60.80  | 77.0      |
| 61.40  | 124.0     |
| 62.70  | 53.0      |
| 63.50  | 56.0      |
| 64.80  | 108.0     |
| 66.00  | 77.0      |
| 67.10  | 115.0     |
| 68.50  | 165.0     |
| 69.20  | 71.0      |
| 70.10  | 88.0      |
| 70.60  | 91.0      |
| 71.10  | 64.0      |
| 71.70  | 58.0      |
| 72.30  | 75.0      |
| 73.20  | 119.0     |
| 73.80  | 52.0      |
| 74.60  | 70.0      |
| 74.90  | 72.0      |
| 75.30  | 56.0      |
| 77.40  | 73.0      |
| 78.00  | 82.0      |
| 79.10  | 68.0      |
| 80.70  | 99.0      |
| 82.50  | 135.0     |
| 83.20  | 72.0      |
| 84.30  | 52.0      |
| 85.00  | 58.0      |
| 86.00  | 84.0      |
| 90.90  | 80.0      |
| 92.40  | 56.0      |
| 92.90  | 57.0      |
| 94.80  | 57.0      |
| 95.00  | 58.0      |
| 98.00  | 89.0      |
| 100.60 | 76.0      |
| 107.00 | 66.0      |
| 108.80 | 75.0      |
| 109.90 | 54.0      |
| 110.80 | 80.0      |
| 111.80 | 52.0      |
| 116.20 | 63.0      |
| 117.20 | 82.0      |
| 118.40 | 56.0      |
| 121.10 | 72.0      |
| 125.10 | 69.0      |
| 127.80 | 150.0     |
| 129.70 | 71.0      |
| 130.70 | 78.0      |
| 131.30 | 67.0      |
| 133.80 | 56.0      |
| 141.50 | 62.0      |
| 144.70 | 66.0      |
| 150.10 | 68.0      |
| 160.10 | 67.0      |
| 161.70 | 87.0      |
| 169.40 | 84.0      |
| 189.30 | 69.0      |
| 190.30 | 62.0      |
| 207.10 | 78.0      |
| 207.30 | 79.0      |
| 228.40 | 68.0      |
| 244.20 | 61.0      |

Instrument :  
MSVOA\_W  
ClientSampleId :  
BFB575

|        |      |
|--------|------|
| 249.70 | 92.0 |
| 255.40 | 57.0 |
| 260.90 | 54.0 |
| 268.60 | 88.0 |
| 272.10 | 53.0 |
| 274.60 | 79.0 |
| 281.10 | 86.0 |
| 293.20 | 52.0 |

**Instrument :**  
MSVOA\_W  
**ClientSampleId :**  
BFB575

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 3301.0    |
| 37.00  | 16322.0   |
| 38.10  | 13837.0   |
| 39.10  | 5595.0    |
| 40.00  | 692.0     |
| 41.00  | 226.0     |
| 42.20  | 133.0     |
| 43.00  | 478.0     |
| 44.10  | 1789.0    |
| 45.00  | 3512.0    |
| 47.10  | 4238.0    |
| 48.00  | 2068.0    |
| 49.10  | 13996.0   |
| 50.10  | 66480.0   |
| 51.10  | 22568.0   |
| 52.00  | 980.0     |
| 52.90  | 324.0     |
| 54.00  | 240.0     |
| 55.00  | 966.0     |
| 56.10  | 5139.0    |
| 57.10  | 9621.0    |
| 58.10  | 504.0     |
| 58.80  | 401.0     |
| 60.00  | 3796.0    |
| 61.00  | 17304.0   |
| 62.00  | 15837.0   |
| 63.00  | 14449.0   |
| 64.00  | 1746.0    |
| 65.00  | 1798.0    |
| 66.00  | 353.0     |
| 68.00  | 37200.0   |
| 69.00  | 36024.0   |
| 70.00  | 2859.0    |
| 71.20  | 305.0     |
| 72.00  | 2088.0    |
| 73.00  | 16752.0   |
| 74.00  | 60520.0   |
| 75.00  | 178624.0  |
| 76.10  | 15283.0   |
| 77.00  | 2309.0    |
| 78.10  | 1374.0    |
| 78.90  | 10360.0   |
| 80.00  | 2930.0    |
| 81.00  | 10948.0   |
| 82.00  | 2044.0    |
| 82.80  | 274.0     |
| 83.40  | 205.0     |
| 83.80  | 241.0     |
| 85.90  | 569.0     |
| 87.00  | 16888.0   |
| 88.00  | 14173.0   |
| 89.60  | 125.0     |
| 91.00  | 1294.0    |
| 92.00  | 9466.0    |
| 93.00  | 15345.0   |
| 94.00  | 39528.0   |
| 95.10  | 337088.0  |
| 96.00  | 25456.0   |
| 97.10  | 829.0     |
| 98.70  | 73.0      |
| 99.60  | 104.0     |
| 101.10 | 90.0      |
| 102.10 | 88.0      |
| 103.00 | 619.0     |
| 104.00 | 1405.0    |
| 105.00 | 696.0     |
| 106.00 | 1316.0    |
| 106.90 | 434.0     |
| 108.00 | 169.0     |
| 110.00 | 331.0     |
| 110.90 | 395.0     |
| 111.60 | 198.0     |
| 112.10 | 385.0     |
| 112.70 | 215.0     |
| 115.00 | 517.0     |
| 116.00 | 1243.0    |
| 116.90 | 1956.0    |
| 117.90 | 1410.0    |
| 118.90 | 2032.0    |
| 119.90 | 168.0     |
| 121.90 | 199.0     |
| 123.00 | 158.0     |
| 123.90 | 251.0     |
| 125.10 | 782.0     |
| 127.10 | 201.0     |
| 128.00 | 1292.0    |
| 129.00 | 848.0     |
| 129.90 | 1500.0    |
| 131.00 | 560.0     |
| 131.80 | 73.0      |
| 132.20 | 108.0     |
| 133.00 | 2037.0    |

Instrument :  
MSVOA\_W  
ClientSampleId :  
BFB575

|        |          |
|--------|----------|
| 133.90 | 315.0    |
| 135.00 | 732.0    |
| 135.90 | 125.0    |
| 136.90 | 493.0    |
| 137.70 | 51.0     |
| 138.30 | 96.0     |
| 139.10 | 216.0    |
| 139.60 | 220.0    |
| 140.90 | 2144.0   |
| 143.00 | 3559.0   |
| 144.00 | 261.0    |
| 144.90 | 1047.0   |
| 146.00 | 482.0    |
| 146.90 | 477.0    |
| 148.00 | 678.0    |
| 149.00 | 325.0    |
| 150.00 | 484.0    |
| 152.00 | 338.0    |
| 152.90 | 307.0    |
| 153.80 | 236.0    |
| 154.90 | 759.0    |
| 155.80 | 135.0    |
| 157.00 | 558.0    |
| 157.90 | 99.0     |
| 159.00 | 339.0    |
| 161.10 | 580.0    |
| 161.90 | 123.0    |
| 163.00 | 321.0    |
| 164.00 | 114.0    |
| 164.90 | 388.0    |
| 165.70 | 63.0     |
| 168.30 | 65.0     |
| 169.10 | 73.0     |
| 169.90 | 143.0    |
| 170.50 | 259.0    |
| 171.20 | 479.0    |
| 171.90 | 3122.0   |
| 173.90 | 228736.0 |
| 175.00 | 18488.0  |
| 176.00 | 240640.0 |
| 176.90 | 16279.0  |
| 178.10 | 518.0    |
| 178.90 | 454.0    |
| 179.80 | 179.0    |
| 180.90 | 107.0    |
| 181.90 | 57.0     |
| 188.90 | 244.0    |
| 191.00 | 1516.0   |
| 192.10 | 426.0    |
| 193.00 | 1953.0   |
| 194.00 | 297.0    |
| 194.80 | 225.0    |
| 196.30 | 86.0     |
| 203.00 | 249.0    |
| 204.20 | 156.0    |
| 205.10 | 459.0    |
| 206.00 | 165.0    |
| 207.00 | 1598.0   |
| 208.10 | 306.0    |
| 208.90 | 288.0    |
| 209.80 | 73.0     |
| 216.70 | 67.0     |
| 219.20 | 75.0     |
| 220.20 | 87.0     |
| 221.00 | 50.0     |
| 234.90 | 425.0    |
| 235.90 | 104.0    |
| 247.90 | 121.0    |
| 248.90 | 1401.0   |
| 249.90 | 303.0    |
| 250.80 | 572.0    |
| 252.00 | 362.0    |
| 253.00 | 163.0    |
| 257.50 | 55.0     |
| 262.00 | 76.0     |
| 265.00 | 2980.0   |
| 266.10 | 777.0    |
| 267.10 | 605.0    |
| 268.00 | 209.0    |
| 269.00 | 84.0     |
| 279.00 | 169.0    |
| 281.10 | 24792.0  |
| 282.10 | 8392.0   |
| 283.00 | 4874.0   |
| 284.10 | 998.0    |
| 285.00 | 331.0    |
| 286.20 | 82.0     |
| 295.20 | 89.0     |

Instrument :  
MSVOA\_W  
ClientSampleId :  
BFB575

| m/z    | Abundance |
|--------|-----------|
| 35.10  | 242.0     |
| 36.00  | 3255.0    |
| 37.10  | 15747.0   |
| 38.10  | 14625.0   |
| 39.10  | 6612.0    |
| 40.10  | 853.0     |
| 40.90  | 122.0     |
| 42.20  | 106.0     |
| 43.10  | 397.0     |
| 44.10  | 2044.0    |
| 45.00  | 3282.0    |
| 47.10  | 3872.0    |
| 48.00  | 2501.0    |
| 49.10  | 14770.0   |
| 50.10  | 68144.0   |
| 51.00  | 20688.0   |
| 52.10  | 885.0     |
| 53.00  | 293.0     |
| 53.60  | 146.0     |
| 54.00  | 166.0     |
| 55.00  | 1090.0    |
| 56.10  | 5719.0    |
| 57.00  | 9778.0    |
| 58.10  | 534.0     |
| 58.80  | 379.0     |
| 60.00  | 3475.0    |
| 61.00  | 17464.0   |
| 62.00  | 16872.0   |
| 63.10  | 13298.0   |
| 64.10  | 1668.0    |
| 65.00  | 1582.0    |
| 66.00  | 192.0     |
| 67.10  | 1318.0    |
| 68.00  | 38440.0   |
| 69.10  | 37168.0   |
| 70.00  | 3326.0    |
| 71.10  | 473.0     |
| 72.00  | 2422.0    |
| 73.10  | 17968.0   |
| 74.10  | 61368.0   |
| 75.00  | 193152.0  |
| 76.10  | 15898.0   |
| 77.00  | 2533.0    |
| 78.90  | 9661.0    |
| 79.90  | 2881.0    |
| 81.00  | 10568.0   |
| 82.00  | 2513.0    |
| 82.80  | 249.0     |
| 83.20  | 206.0     |
| 83.70  | 134.0     |
| 85.20  | 114.0     |
| 87.00  | 18144.0   |
| 88.00  | 16253.0   |
| 89.70  | 87.0      |
| 91.00  | 1548.0    |
| 92.10  | 9896.0    |
| 93.00  | 16416.0   |
| 94.10  | 45528.0   |
| 95.00  | 401216.0  |
| 96.00  | 25024.0   |
| 97.10  | 1002.0    |
| 97.80  | 67.0      |
| 98.70  | 67.0      |
| 101.10 | 117.0     |
| 101.80 | 88.0      |
| 102.60 | 230.0     |
| 103.20 | 283.0     |
| 104.00 | 1453.0    |
| 104.90 | 776.0     |
| 105.90 | 1433.0    |
| 107.00 | 435.0     |
| 108.20 | 84.0      |
| 109.80 | 280.0     |
| 110.90 | 484.0     |
| 112.00 | 415.0     |
| 112.70 | 282.0     |
| 113.20 | 242.0     |
| 115.00 | 445.0     |
| 116.00 | 1522.0    |
| 117.00 | 2210.0    |
| 118.00 | 1229.0    |
| 118.90 | 2106.0    |
| 121.10 | 60.0      |
| 121.50 | 78.0      |
| 122.10 | 124.0     |
| 122.80 | 184.0     |
| 123.90 | 267.0     |
| 124.90 | 518.0     |
| 125.80 | 371.0     |
| 126.20 | 264.0     |
| 127.00 | 244.0     |
| 127.90 | 943.0     |

Instrument :  
MSVOA\_W  
ClientSampleId :  
BFB575

|        |          |
|--------|----------|
| 128.90 | 836.0    |
| 130.00 | 1687.0   |
| 131.00 | 627.0    |
| 131.70 | 59.0     |
| 132.10 | 135.0    |
| 133.00 | 1295.0   |
| 133.90 | 268.0    |
| 134.30 | 369.0    |
| 134.90 | 902.0    |
| 136.00 | 212.0    |
| 136.20 | 211.0    |
| 136.90 | 572.0    |
| 137.80 | 58.0     |
| 138.30 | 132.0    |
| 138.90 | 163.0    |
| 139.10 | 163.0    |
| 139.90 | 223.0    |
| 140.90 | 3570.0   |
| 142.10 | 556.0    |
| 142.90 | 3795.0   |
| 144.00 | 326.0    |
| 145.00 | 724.0    |
| 145.90 | 438.0    |
| 146.90 | 357.0    |
| 147.90 | 598.0    |
| 148.80 | 316.0    |
| 150.00 | 296.0    |
| 150.90 | 95.0     |
| 151.90 | 267.0    |
| 152.90 | 326.0    |
| 153.90 | 486.0    |
| 155.00 | 770.0    |
| 156.20 | 162.0    |
| 157.00 | 729.0    |
| 158.00 | 104.0    |
| 158.90 | 534.0    |
| 160.90 | 624.0    |
| 161.90 | 112.0    |
| 163.00 | 301.0    |
| 164.00 | 94.0     |
| 164.90 | 220.0    |
| 166.00 | 79.0     |
| 167.40 | 54.0     |
| 168.70 | 102.0    |
| 169.70 | 97.0     |
| 170.90 | 361.0    |
| 172.00 | 2805.0   |
| 173.90 | 306304.0 |
| 175.00 | 23288.0  |
| 176.00 | 293504.0 |
| 177.00 | 19912.0  |
| 178.00 | 711.0    |
| 178.80 | 271.0    |
| 180.10 | 120.0    |
| 180.80 | 76.0     |
| 184.90 | 66.0     |
| 187.60 | 104.0    |
| 188.70 | 87.0     |
| 189.60 | 126.0    |
| 191.10 | 1050.0   |
| 192.30 | 271.0    |
| 193.00 | 952.0    |
| 193.80 | 213.0    |
| 194.30 | 184.0    |
| 195.60 | 55.0     |
| 202.70 | 54.0     |
| 203.20 | 85.0     |
| 204.90 | 401.0    |
| 207.00 | 963.0    |
| 208.10 | 217.0    |
| 209.30 | 99.0     |
| 210.40 | 65.0     |
| 219.10 | 59.0     |
| 220.30 | 56.0     |
| 221.10 | 54.0     |
| 222.50 | 72.0     |
| 230.90 | 80.0     |
| 234.40 | 81.0     |
| 234.90 | 392.0    |
| 236.90 | 97.0     |
| 249.10 | 950.0    |
| 250.00 | 338.0    |
| 251.00 | 524.0    |
| 252.10 | 123.0    |
| 253.10 | 132.0    |
| 259.10 | 88.0     |
| 262.60 | 81.0     |
| 265.10 | 1410.0   |
| 266.10 | 477.0    |
| 267.00 | 465.0    |
| 268.10 | 153.0    |
| 268.90 | 52.0     |
| 277.10 | 71.0     |
| 278.90 | 92.0     |
| 281.00 | 19448.0  |
| 282.10 | 5636.0   |

Instrument :  
MSVOA\_W  
ClientSampleId :  
BFB575

|        |        |
|--------|--------|
| 283.00 | 3052.0 |
| 284.00 | 889.0  |
| 284.90 | 281.0  |
| 286.00 | 74.0   |

**Instrument :**  
MSVOA\_W  
**ClientSampleId :**  
BFB575



| m/z    | Abundance |
|--------|-----------|
| 35.10  | 266.0     |
| 36.10  | 2746.0    |
| 37.10  | 10895.0   |
| 38.00  | 11001.0   |
| 39.10  | 4351.0    |
| 40.00  | 658.0     |
| 40.70  | 185.0     |
| 42.00  | 181.0     |
| 43.10  | 290.0     |
| 44.10  | 1988.0    |
| 45.00  | 2506.0    |
| 47.00  | 3757.0    |
| 48.10  | 1725.0    |
| 49.10  | 10110.0   |
| 50.10  | 51120.0   |
| 51.10  | 16249.0   |
| 52.10  | 584.0     |
| 53.00  | 164.0     |
| 53.60  | 137.0     |
| 54.90  | 856.0     |
| 56.00  | 3936.0    |
| 57.10  | 7083.0    |
| 58.10  | 388.0     |
| 58.70  | 229.0     |
| 59.20  | 269.0     |
| 60.00  | 2779.0    |
| 61.00  | 13862.0   |
| 62.00  | 12527.0   |
| 63.10  | 9921.0    |
| 64.10  | 1368.0    |
| 65.00  | 1050.0    |
| 66.00  | 306.0     |
| 67.10  | 987.0     |
| 68.00  | 28088.0   |
| 69.10  | 27432.0   |
| 70.10  | 2258.0    |
| 71.00  | 338.0     |
| 72.10  | 1569.0    |
| 73.00  | 12603.0   |
| 74.10  | 49000.0   |
| 75.10  | 147648.0  |
| 76.10  | 11931.0   |
| 77.00  | 1910.0    |
| 77.90  | 1549.0    |
| 78.90  | 8343.0    |
| 80.00  | 2586.0    |
| 81.00  | 7909.0    |
| 82.00  | 1636.0    |
| 82.90  | 403.0     |
| 85.00  | 143.0     |
| 86.00  | 473.0     |
| 87.00  | 13626.0   |
| 88.00  | 12957.0   |
| 89.90  | 210.0     |
| 91.10  | 977.0     |
| 92.00  | 7743.0    |
| 93.00  | 11423.0   |
| 94.00  | 34488.0   |
| 95.00  | 309248.0  |
| 96.00  | 19776.0   |
| 97.10  | 599.0     |
| 97.90  | 122.0     |
| 99.30  | 51.0      |
| 101.40 | 58.0      |
| 101.80 | 68.0      |
| 103.00 | 336.0     |
| 104.00 | 850.0     |
| 104.80 | 474.0     |
| 105.90 | 1256.0    |
| 106.80 | 245.0     |
| 107.20 | 191.0     |
| 108.00 | 124.0     |
| 109.30 | 150.0     |
| 110.00 | 368.0     |
| 111.00 | 316.0     |
| 111.70 | 163.0     |
| 112.20 | 231.0     |
| 112.90 | 219.0     |
| 114.80 | 377.0     |
| 115.80 | 781.0     |
| 117.00 | 1914.0    |
| 117.80 | 1153.0    |
| 119.00 | 1292.0    |
| 122.00 | 123.0     |
| 122.90 | 143.0     |
| 123.80 | 141.0     |
| 125.10 | 191.0     |
| 125.50 | 266.0     |
| 126.50 | 104.0     |
| 128.00 | 985.0     |
| 128.90 | 660.0     |
| 130.00 | 901.0     |

Instrument :  
MSVOA\_W  
ClientSampleId :  
BFB575

|        |          |
|--------|----------|
| 131.00 | 439.0    |
| 132.90 | 685.0    |
| 133.90 | 202.0    |
| 135.00 | 517.0    |
| 135.80 | 88.0     |
| 136.90 | 616.0    |
| 137.80 | 115.0    |
| 138.90 | 164.0    |
| 139.50 | 169.0    |
| 141.00 | 2186.0   |
| 142.00 | 369.0    |
| 142.90 | 2938.0   |
| 144.00 | 334.0    |
| 145.00 | 496.0    |
| 145.90 | 400.0    |
| 147.00 | 371.0    |
| 147.80 | 744.0    |
| 148.90 | 341.0    |
| 149.90 | 257.0    |
| 152.00 | 167.0    |
| 152.80 | 221.0    |
| 154.10 | 216.0    |
| 155.00 | 798.0    |
| 156.90 | 658.0    |
| 158.90 | 394.0    |
| 160.90 | 390.0    |
| 162.00 | 87.0     |
| 162.80 | 119.0    |
| 163.40 | 91.0     |
| 165.00 | 97.0     |
| 169.70 | 118.0    |
| 170.20 | 90.0     |
| 170.80 | 361.0    |
| 172.00 | 1883.0   |
| 173.90 | 246208.0 |
| 176.00 | 233472.0 |
| 176.90 | 16744.0  |
| 177.90 | 522.0    |
| 178.70 | 245.0    |
| 180.10 | 51.0     |
| 184.90 | 51.0     |
| 189.60 | 113.0    |
| 191.10 | 431.0    |
| 192.00 | 164.0    |
| 193.00 | 435.0    |
| 194.00 | 164.0    |
| 194.70 | 98.0     |
| 195.50 | 53.0     |
| 205.00 | 259.0    |
| 207.00 | 582.0    |
| 208.20 | 98.0     |
| 209.00 | 76.0     |
| 210.50 | 51.0     |
| 222.60 | 92.0     |
| 231.00 | 60.0     |
| 233.10 | 84.0     |
| 234.80 | 165.0    |
| 236.90 | 55.0     |
| 248.00 | 50.0     |
| 248.90 | 408.0    |
| 249.90 | 192.0    |
| 251.20 | 315.0    |
| 262.60 | 81.0     |
| 265.10 | 789.0    |
| 265.70 | 248.0    |
| 266.20 | 291.0    |
| 266.90 | 243.0    |
| 267.80 | 75.0     |
| 277.10 | 52.0     |
| 279.70 | 111.0    |
| 281.10 | 6692.0   |
| 282.10 | 1592.0   |
| 283.00 | 1220.0   |
| 284.20 | 362.0    |
| 284.70 | 66.0     |
| 286.00 | 82.0     |

Instrument :  
MSVOA\_W  
ClientSampleId :  
BFB575