

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090319\
 Data File : VW012547.D
 Acq On : 03 Sep 2019 15:30
 Operator : SY/VA
 Sample : K4544-03
 Misc : 3.80G/5ML/MSVOA W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T16-17-B2-(0)

Quant Time: Sep 03 16:18:52 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	221539	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	352907	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	266280	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	77184	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	134892	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
35) Dibromofluoromethane	7.88	113	109949	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
50) Toluene-d8	10.32	98	418448	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
62) 4-Bromofluorobenzene	12.62	95	106653	35.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.84%	
Target Compounds						
20) Methylene Chloride	4.92	84	12281	4.826	ug/l	92
29) Tetrahydrofuran	7.54	42	5353	9.753	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090319\
Data File : VW012547.D
Acq On : 03 Sep 2019 15:30
Operator : SY/VA
Sample : K4544-03
Misc : 3.80G/5ML/MSVOA W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T16-17-B2-(0)

Quant Time: Sep 03 16:18:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

