

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090319\
 Data File : VW012548.D
 Acq On : 03 Sep 2019 15:56
 Operator : SY/VA
 Sample : K4514-05
 Misc : 5.39G/5ML/MSVOA W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S602-M-2.0-2.5-082319

Quant Time: Sep 21 01:32:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	212178	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	334776	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	217577	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	47180	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	133406	52.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.78%	
35) Dibromofluoromethane	7.88	113	106950	53.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.04%	
50) Toluene-d8	10.32	98	385137	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
62) 4-Bromofluorobenzene	12.62	95	75797	26.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	53.08%	
Target Compounds						
4) Vinyl Chloride	2.35	62	3904	1.576	ug/l	91
16) Acetone	4.12	43	151459	234.912	ug/l	99
17) Carbon Disulfide	4.38	76	9298	1.446	ug/l	95
25) 2-Butanone	7.18	43	30681	35.018	ug/l	93
27) cis-1,2-Dichloroethene	7.17	96	6139	2.518	ug/l	90
29) Tetrahydrofuran	7.54	42	5250	9.988	ug/l	98
52) Toluene	10.38	92	13176	2.322	ug/l	98
67) Ethyl Benzene	11.73	91	10658	1.291	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090319\
Data File : VW012548.D
Acq On : 03 Sep 2019 15:56
Operator : SY/VA
Sample : K4514-05
Misc : 5.39G/5ML/MSVOA W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S602-M-2.0-2.5-082319

Quant Time: Sep 21 01:32:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

