

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090319\
 Data File : VW012568.D
 Acq On : 04 Sep 2019 01:49
 Operator : SY/VA
 Sample : K4544-17
 Misc : 4.41G/5ML/MSVOA W/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T16-17-C3-(2.25)

Quant Time: Sep 04 08:00:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	26347	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	40457	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	32775	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	7532	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	18101	57.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.58%	
35) Dibromofluoromethane	7.88	113	13863	57.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.82%	
50) Toluene-d8	10.32	98	47846	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
62) 4-Bromofluorobenzene	12.62	95	12193	35.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.66%	
Target Compounds						
16) Acetone	4.13	43	3911	48.850	ug/l	93
20) Methylene Chloride	4.92	84	3650	12.060	ug/l #	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090319\
Data File : VW012568.D
Acq On : 04 Sep 2019 01:49
Operator : SY/VA
Sample : K4544-17
Misc : 4.41G/5ML/MSVOA W/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T16-17-C3-(2.25)

Quant Time: Sep 04 08:00:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

