

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090319\
 Data File : VW012577.D
 Acq On : 04 Sep 2019 05:41
 Operator : SY/VA
 Sample : K4525-12
 Misc : 4.91G/5ML/MSVOA W/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T9-12-13-B2-(0)

Quant Time: Sep 04 07:49:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	263582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	424651	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	350413	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	118650	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	174294	55.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.26%	
35) Dibromofluoromethane	7.88	113	132884	52.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.84%	
50) Toluene-d8	10.32	98	529971	51.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.14%	
62) 4-Bromofluorobenzene	12.62	95	152400	42.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.14%	
Target Compounds						
16) Acetone	4.14	43	31493	39.320	ug/l	95
20) Methylene Chloride	4.90	84	25041	8.270	ug/l	94
29) Tetrahydrofuran	7.54	42	5451	8.348	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090319\
Data File : VW012577.D
Acq On : 04 Sep 2019 05:41
Operator : SY/VA
Sample : K4525-12
Misc : 4.91G/5ML/MSVOA W/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T9-12-13-B2-(0)

Quant Time: Sep 04 07:49:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

