

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090319\
 Data File : VW012579.D
 Acq On : 04 Sep 2019 06:32
 Operator : SY/VA
 Sample : K4525-14
 Misc : 5.49G/5ML/MSVOA W/SOIL
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T9-12-13-B4-(0)

Quant Time: Sep 04 07:39:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	255882	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	417017	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	348868	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	135983	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	170715	56.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.24%	
35) Dibromofluoromethane	7.88	113	131054	52.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.30%	
50) Toluene-d8	10.32	98	510136	50.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.08%	
62) 4-Bromofluorobenzene	12.62	95	156680	44.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.08%	
Target Compounds						
16) Acetone	4.14	43	21762	27.988	ug/l	92
20) Methylene Chloride	4.90	84	36404	12.385	ug/l	99
29) Tetrahydrofuran	7.53	42	5893	9.296	ug/l	94
52) Toluene	10.38	92	19707	2.788	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090319\
Data File : VW012579.D
Acq On : 04 Sep 2019 06:32
Operator : SY/VA
Sample : K4525-14
Misc : 5.49G/5ML/MSVOA W/SOIL
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T9-12-13-B4-(0)

Quant Time: Sep 04 07:39:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

