

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090418\
 Data File : VW005116.D
 Acq On : 05 Sep 2018 08:31
 Operator : SY/AP
 Sample : J4712-08
 Misc : 5.06G/5ML/MSVOA W/SOIL
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-07

Quant Time: Sep 05 10:56:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	585929	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	953890	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	833583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	315292	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	189211	38.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.92%	
35) Dibromofluoromethane	7.89	113	185229	33.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.20%	
50) Toluene-d8	10.33	98	660946	26.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	52.04%	
62) 4-Bromofluorobenzene	12.62	95	240309	26.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	53.66%	
Target Compounds						
8) Diethyl Ether	3.69	74	5311	1.96	ug/l	Qvalue 84
20) Methylene Chloride	4.92	84	81619	9.23	ug/l	94
52) Toluene	10.40	92	23184	1.23	ug/l	97
84) 1,2,4-Trimethylbenzene	13.26	105	20868	1.07	ug/l	100
85) sec-Butylbenzene	13.39	105	34730	1.47	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090418\
Data File : VW005116.D
Acq On : 05 Sep 2018 08:31
Operator : SY/AP
Sample : J4712-08
Misc : 5.06G/5ML/MSVOA W/SOIL
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-07

Quant Time: Sep 05 10:56:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

