

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090518\
 Data File : VW005147.D
 Acq On : 05 Sep 2018 23:00
 Operator : SY/AP
 Sample : J4712-01RE
 Misc : 5.78G/5ML/MSVOA W/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-13RE

Quant Time: Sep 06 03:59:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	559638	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	874967	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	829489	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	320705	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	294951	63.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.16%	
35) Dibromofluoromethane	7.89	113	266989	52.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.60%	
50) Toluene-d8	10.33	98	1212347	52.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.08%	
62) 4-Bromofluorobenzene	12.62	95	372647	45.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.72%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.92	84	71544	8.12	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090518\
Data File : VW005147.D
Acq On : 05 Sep 2018 23:00
Operator : SY/AP
Sample : J4712-01RE
Misc : 5.78G/5ML/MSVOA W/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-13RE

Quant Time: Sep 06 03:59:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

