


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW090518\
Data File : VW005127.D
Acq On : 05 Sep 2018 13:19
Operator : SY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Title : SW846 8260
Last Update : Sat Sep 01 03:35:05 2018



AutoFind: Scans 1809, 1810, 1811; Background Corrected with Scan 1795

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	61788	PASS
75	95	30	60	46.7	172394	PASS
95	95	100	100	100.0	369105	PASS
96	95	5	9	6.9	25507	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	79.8	294642	PASS
175	174	5	9	7.5	22151	PASS
176	174	95	101	95.8	282261	PASS
177	176	5	9	7.1	20084	PASS