

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090618\
 Data File : VW005165.D
 Acq On : 06 Sep 2018 15:51
 Operator : SY/AP
 Sample : J4720-04RE
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BZTRANS-4RE

Quant Time: Sep 06 16:50:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	378862	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	600444	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	480339	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	158142	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.32	65	227686	72.50	ug/l	0.00
Spiked Amount						
			Recovery	=	145.00%	
35) Dibromofluoromethane	7.89	113	203472	58.64	ug/l	0.00
Spiked Amount						
			Recovery	=	117.28%	
50) Toluene-d8	10.33	98	711556	44.50	ug/l	0.00
Spiked Amount						
			Recovery	=	89.00%	
62) 4-Bromofluorobenzene	12.63	95	198668	35.24	ug/l	0.00
Spiked Amount						
			Recovery	=	70.48%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.14	43	5704	9.06	ug/l	97
20) Methylene Chloride	4.92	84	66763	12.79	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090618\
Data File : VW005165.D
Acq On : 06 Sep 2018 15:51
Operator : SY/AP
Sample : J4720-04RE
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BZTRANS-4RE

Quant Time: Sep 06 16:50:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

