

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090618\
 Data File : VW005190.D
 Acq On : 07 Sep 2018 04:39
 Operator : SY/AP
 Sample : J4693-04RE
 Misc : 6.29G/5ML/MSVOA W/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EP1-MW-F(7-10)RE

Quant Time: Sep 07 14:28:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	375425	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	678276	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	473905	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	570593	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	252504	81.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	162.28%	
35) Dibromofluoromethane	7.89	113	246645	62.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.86%	
50) Toluene-d8	10.33	98	723311	40.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.10%	
62) 4-Bromofluorobenzene	12.61	95	17500811	2748.03	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	5496.06%	
Target Compounds						
16) Acetone	4.14	43	33955	54.40	ug/l	93
17) Carbon Disulfide	4.38	76	265719	25.12	ug/l	100
20) Methylene Chloride	4.92	84	32343	4.10	ug/l	96
85) sec-Butylbenzene	13.40	105	7418051	173.72	ug/l #	53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090618\
Data File : VW005190.D
Acq On : 07 Sep 2018 04:39
Operator : SY/AP
Sample : J4693-04RE
Misc : 6.29G/5ML/MSVOA W/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EP1-MW-F(7-10)RE

Quant Time: Sep 07 14:28:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

