

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090618\
 Data File : VW005200.D
 Acq On : 07 Sep 2018 08:55
 Operator : SY/AP
 Sample : J4782-01
 Misc : 6.79G/5ML/MSVOA W/SOIL
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ENY-SB-03-60.5-61

Quant Time: Sep 07 09:16:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	425165	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	735459	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	747481	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	373129	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	239973	68.09	ug/l	0.00
Spiked Amount			Recovery	=	136.18%	
35) Dibromofluoromethane	7.89	113	215919	50.80	ug/l	0.00
Spiked Amount			Recovery	=	101.60%	
50) Toluene-d8	10.33	98	971297	49.60	ug/l	0.00
Spiked Amount			Recovery	=	99.20%	
62) 4-Bromofluorobenzene	12.62	95	349351	50.59	ug/l	0.00
Spiked Amount			Recovery	=	101.18%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.14	43	37722	53.37	ug/l	95
20) Methylene Chloride	4.92	84	53382	7.90	ug/l	96
30) Chloroform	7.68	83	13607	1.89	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090618\
Data File : VW005200.D
Acq On : 07 Sep 2018 08:55
Operator : SY/AP
Sample : J4782-01
Misc : 6.79G/5ML/MSVOA W/SOIL
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ENY-SB-03-60.5-61

Quant Time: Sep 07 09:16:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

