

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090718\
 Data File : VW005206.D
 Acq On : 07 Sep 2018 14:00
 Operator : SY/AP
 Sample : J4720-02
 Misc : 5.10G/5ML/MSVOA W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BZTRANS-2

Quant Time: Sep 08 00:18:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	533588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	814711	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	685537	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	251044	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	195075	44.10	ug/l	0.00
Spiked Amount			Recovery	=	88.20%	
35) Dibromofluoromethane	7.89	113	186824	39.68	ug/l	0.00
Spiked Amount			Recovery	=	79.36%	
50) Toluene-d8	10.33	98	755578	34.83	ug/l	0.00
Spiked Amount			Recovery	=	69.66%	
62) 4-Bromofluorobenzene	12.63	95	213474	27.91	ug/l	0.00
Spiked Amount			Recovery	=	55.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.92	84	65277	7.59	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090718\
Data File : VW005206.D
Acq On : 07 Sep 2018 14:00
Operator : SY/AP
Sample : J4720-02
Misc : 5.10G/5ML/MSVOA W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BZTRANS-2

Quant Time: Sep 08 00:18:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

