

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090718\
 Data File : VW005212.D
 Acq On : 07 Sep 2018 16:33
 Operator : SY/AP
 Sample : J4748-03
 Misc : 5.55G/5ML/MSVOA W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-08

Quant Time: Sep 08 01:40:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	499007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	776541	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	771338	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	376798	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	214070	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
35) Dibromofluoromethane	7.89	113	196605	43.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.62%	
50) Toluene-d8	10.33	98	911913	44.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.20%	
62) 4-Bromofluorobenzene	12.63	95	312975	42.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.86%	
Target Compounds						
8) Diethyl Ether	3.70	74	6319	2.74	ug/l	96
16) Acetone	4.13	43	26341	31.75	ug/l	94
20) Methylene Chloride	4.92	84	82866	11.81	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090718\
Data File : VW005212.D
Acq On : 07 Sep 2018 16:33
Operator : SY/AP
Sample : J4748-03
Misc : 5.55G/5ML/MSVOA W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-08

Quant Time: Sep 08 01:40:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

