

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090718\
 Data File : VW005215.D
 Acq On : 07 Sep 2018 17:50
 Operator : SY/AP
 Sample : J4748-06RE
 Misc : 6.51G/5ML/MSVOA W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DW-11RE

Manual Integrations
 APPROVED

MMDadoda
 9/10/2018 2:52:33 PM

Quant Time: Sep 10 12:16:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 08 02:36:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	419969	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	679088	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	458450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	139312	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	131274	37.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.42%	
35) Dibromofluoromethane	7.89	113	116716	29.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	59.48%	
50) Toluene-d8	10.33	98	339289	18.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	37.52%	
62) 4-Bromofluorobenzene	12.63	95	106362	16.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	33.36%	
Target Compounds						
17) Carbon Disulfide	4.38	76	36826	3.11	ug/l	Qvalue 100
18) Methyl Acetate	4.68	43	4819m	2.35	ug/l	
20) Methylene Chloride	4.92	84	51662	9.06	ug/l	96
52) Toluene	10.39	92	14040	1.05	ug/l	94
86) p-Isopropyltoluene	13.51	119	17261	1.87	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090718\
Data File : VW005215.D
Acq On : 07 Sep 2018 17:50
Operator : SY/AP
Sample : J4748-06RE
Misc : 6.51G/5ML/MSVOA W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DW-11RE

Manual Integrations
APPROVED

MMDadoda
9/10/2018 2:52:33 PM

Quant Time: Sep 10 12:16:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 08 02:36:58 2018
Response via : Initial Calibration

