

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090718\
 Data File : VW005220.D
 Acq On : 07 Sep 2018 19:57
 Operator : SY/AP
 Sample : J4775-03
 Misc : 6.36G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 OR-02-090618-B

Quant Time: Sep 08 02:01:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 01 03:35:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	473323	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	720693	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	552569	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	220592	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	167707	42.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.48%	
35) Dibromofluoromethane	7.89	113	154579	37.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.24%	
50) Toluene-d8	10.33	98	534088	27.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	55.66%	
62) 4-Bromofluorobenzene	12.62	95	185150	27.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	54.72%	
Target Compounds						
16) Acetone	4.13	43	37835	48.08	ug/l	89
69) o-Xylene	12.17	106	21108	2.71	ug/l	100
80) 1,3,5-Trimethylbenzene	12.95	105	29430	2.19	ug/l	98
84) 1,2,4-Trimethylbenzene	13.26	105	16537	1.21	ug/l	99
95) Naphthalene	15.38	128	300094	38.71	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090718\
Data File : VW005220.D
Acq On : 07 Sep 2018 19:57
Operator : SY/AP
Sample : J4775-03
Misc : 6.36G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
OR-02-090618-B

Quant Time: Sep 08 02:01:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090118S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 01 03:35:05 2018
Response via : Initial Calibration

