

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090722\
 Data File : VW024565.D
 Acq On : 07 Sep 2022 11:39
 Operator : SY/VA
 Sample : N4322-20
 Misc : 5.06g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Sep 08 01:19:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Sep 07 04:52:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	302646	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	292640	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	128766	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	161498	25.846	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 103.400%		
7) Chloroethane-d5	2.879	69	124086	26.708	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 106.840%		
11) 1,1-Dichloroethene-d2	4.013	63	124876	18.942	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 75.760%		
21) 2-Butanone-d5	7.074	46	47256	46.881	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 93.760%		
24) Chloroform-d	7.647	84	190962	24.774	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 99.080%		
26) 1,2-Dichloroethane-d4	8.305	65	113630	26.146	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 104.600%		
32) Benzene-d6	8.268	84	392660	27.278	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 109.120%		
36) 1,2-Dichloropropane-d6	9.274	67	109259	25.743	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 102.960%		
41) Toluene-d8	10.323	98	364286	25.959	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 103.840%		
43) trans-1,3-Dichloroprop...	10.573	79	46242	24.313	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 97.240%		
47) 2-Hexanone-d5	10.920	63	39192	50.238	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 100.480%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	93906	23.085	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 92.320%		
66) 1,2-Dichlorobenzene-d4	13.853	152	109831	26.930	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 107.720%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090722\
Data File : VW024565.D
Acq On : 07 Sep 2022 11:39
Operator : SY/VA
Sample : N4322-20
Misc : 5.06g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Sep 08 01:19:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090622SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Sep 07 04:52:14 2022
Response via : Initial Calibration

