

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090722\
 Data File : VW024566.D
 Acq On : 07 Sep 2022 12:03
 Operator : SY/VA
 Sample : N4346-18
 Misc : 5.02g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Sep 08 01:19:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Sep 07 04:52:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	304018	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	294040	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	126697	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	159785	25.457	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.840%
7) Chloroethane-d5	2.879	69	121418	26.015	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	104.080%
11) 1,1-Dichloroethene-d2	4.007	63	122588	18.511	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	74.040%
21) 2-Butanone-d5	6.872	46	237	0.234	ug/L	-0.21
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.460%#
24) Chloroform-d	7.646	84	189146	24.428	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.720%
26) 1,2-Dichloroethane-d4	8.549	65	172	0.039	ug/L	0.24
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.160%#
32) Benzene-d6	8.274	84	384678	26.596	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	106.400%
36) 1,2-Dichloropropane-d6	9.274	67	109565	25.693	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.760%
41) Toluene-d8	10.323	98	359568	25.501	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	102.000%
43) trans-1,3-Dichloroprop...	10.579	79	47127	24.660	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	98.640%
47) 2-Hexanone-d5	10.920	63	40395	51.534	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.060%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	94247	23.058	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	92.240%
66) 1,2-Dichlorobenzene-d4	13.853	152	108079	26.933	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	107.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090722\
Data File : VW024566.D
Acq On : 07 Sep 2022 12:03
Operator : SY/VA
Sample : N4346-18
Misc : 5.02g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Sep 08 01:19:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090622SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Sep 07 04:52:14 2022
Response via : Initial Calibration

