

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090819\
 Data File : VW012771.D
 Acq On : 08 Sep 2019 16:01
 Operator : SY/VA
 Sample : K4603-08
 Misc : 4.46G/5ML/MSVOA W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PAR-E2-(3.11)

Quant Time: Sep 09 07:35:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090719S.M
 Quant Title : SW846 8260
 QLast Update : Sun Sep 08 05:04:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	226388	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	343002	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	291998	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	120651	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	133400	52.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.48%	
35) Dibromofluoromethane	7.88	113	102243	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
50) Toluene-d8	10.32	98	408839	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
62) 4-Bromofluorobenzene	12.62	95	131095	45.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.40%	
Target Compounds						
16) Acetone	4.15	43	28913	41.928	ug/l	95
29) Tetrahydrofuran	7.54	42	4448	7.987	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090819\
Data File : VW012771.D
Acq On : 08 Sep 2019 16:01
Operator : SY/VA
Sample : K4603-08
Misc : 4.46G/5ML/MSVOA W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PAR-E2-(3.11)

Quant Time: Sep 09 07:35:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090719S.M
Quant Title : SW846 8260
QLast Update : Sun Sep 08 05:04:40 2019
Response via : Initial Calibration

