

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090819\
 Data File : VW012803.D
 Acq On : 09 Sep 2019 06:25
 Operator : SY/VA
 Sample : K4623-15
 Misc : 4.61G/5ML/MSVOA W/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 PAR-11-D3-(4)

Quant Time: Sep 09 07:59:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090719S.M
 Quant Title : SW846 8260
 QLast Update : Sun Sep 08 05:04:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	256374	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	385502	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	292733	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	92477	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	146249	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
35) Dibromofluoromethane	7.88	113	115210	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
50) Toluene-d8	10.32	98	453672	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
62) 4-Bromofluorobenzene	12.62	95	118447	36.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.68%	
Target Compounds						
16) Acetone	4.14	43	32423	41.518	ug/l	91
29) Tetrahydrofuran	7.54	42	2970	4.710	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090819\
Data File : VW012803.D
Acq On : 09 Sep 2019 06:25
Operator : SY/VA
Sample : K4623-15
Misc : 4.61G/5ML/MSVOA W/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
PAR-11-D3-(4)

Quant Time: Sep 09 07:59:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090719S.M
Quant Title : SW846 8260
QLast Update : Sun Sep 08 05:04:40 2019
Response via : Initial Calibration

